

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2600

Client ID : SSTDCCC040

Date Analyzed: 07/18/2025

Lab File ID: BP025187.D

Time Analyzed: 11:04

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	531610	7.437	2146080	10.19	1427550	14.10
UPPER LIMIT	1063220	7.937	4292160	10.69	2855100	14.601
LOWER LIMIT	265805	6.937	1073040	9.69	713775	13.601
EPA SAMPLE NO.						
01 CHRT-26899MS	725237	7.44	2872260	10.18	1756650	14.10
02 CHRT-26899MSD	718069	7.44	2875360	10.19	1762350	14.10

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.