

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025154.D
Instrument ID: BNA_P

Contract: TAC001
SDG NO.: Q2600
DFTPP Injection Date: 07/16/2025
DFTPP Injection Time: 11:19

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	28.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	14.5
442	Greater than 50% of mass 198	93.6
443	15.0 - 24.0% of mass 442	18.5 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025155.D	07/16/2025	12:42
STOCK-PILE	Q2600-05	BP025161.D	07/16/2025	16:54
TRENCH	Q2600-01	BP025163.D	07/16/2025	18:17
END-OF-TRENCH	Q2600-09	BP025164.D	07/16/2025	18:59