

Report of Analysis

Client:	T&A Construction Inc		Date Collected:	07/16/25	
Project:	Kingsland Point Park Water Main		Date Received:	07/16/25	
Client Sample ID:	PB168847TB		SDG No.:	Q2600	
Lab Sample ID:	PB168847TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143120.D	1	07/16/25 10:14	07/16/25 21:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	115		23 - 138	77%	SPK: 150
13127-88-3	Phenol-d6	114		10 - 134	76%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.0		67 - 132	85%	SPK: 100
321-60-8	2-Fluorobiphenyl	73.8		52 - 132	74%	SPK: 100
118-79-6	2,4,6-Tribromophenol	134		44 - 137	90%	SPK: 150
1718-51-0	Terphenyl-d14	84.3		42 - 152	84%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	136000	6.963			
1146-65-2	Naphthalene-d8	521000	8.245			
15067-26-2	Acenaphthene-d10	267000	10.004			
1517-22-2	Phenanthrene-d10	451000	11.486			
1719-03-5	Chrysene-d12	236000	14.127			
1520-96-3	Perylene-d12	211000	15.633			

Report of Analysis

Client:	T&A Construction Inc		Date Collected:	07/16/25	
Project:	Kingsland Point Park Water Main		Date Received:	07/16/25	
Client Sample ID:	PB168847TB		SDG No.:	Q2600	
Lab Sample ID:	PB168847TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143120.D	1	07/16/25 10:14	07/16/25 21:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products