

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2600

Analytical Method: SW8270E

Client: T&A Construction Inc

DataFile: BF143123.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2592-02MSD		Client Sample ID: WC-SOIL-20250711MSD									
Pyridine	500	0	290	ug/L	58	0			10	109	20
1,4-Dichlorobenzene	500	0	360	ug/L	72	0			55	125	20
2-Methylphenol	500	0	400	ug/L	80	3			60	131	20
3+4-Methylphenols	500	0	400	ug/L	80	3			54	136	20
Hexachloroethane	500	0	380	ug/L	76	3			19	146	20
Nitrobenzene	500	0	430	ug/L	86	0			62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78	0			52	125	20
2,4,6-Trichlorophenol	500	0	440	ug/L	88	0			78	112	20
2,4,5-Trichlorophenol	500	0	430	ug/L	86	0			71	111	20
2,4-Dinitrotoluene	500	0	470	ug/L	94	2			74	137	20
Hexachlorobenzene	500	0	440	ug/L	88	2			72	115	20
Pentachlorophenol	1000	0	710	ug/L	71	5			52	162	20