

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070225\
 Data File : VX046867.D
 Acq On : 02 Jul 2025 15:14
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX070225

Quant Time: Jul 03 05:55:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.478	0.475	0.6	100	0.00
3 P	Chloromethane	0.522	0.512	1.9	103	0.00
4 C	Vinyl Chloride	0.569	0.555	2.5#	103	0.00
5 T	Bromomethane	0.366	0.375	-2.5	107	0.00
6 T	Chloroethane	0.376	0.351	6.6	104	0.00
7 T	Trichlorofluoromethane	0.880	0.868	1.4	103	0.00
8 T	Diethyl Ether	0.313	0.311	0.6	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.561	0.555	1.1	103	0.00
10 T	Methyl Iodide	0.611	0.611	0.0	99	0.00
11 T	Tert butyl alcohol	0.043	0.046	-7.0	112	0.00
12 CM	1,1-Dichloroethene	0.542	0.526	3.0#	103	0.00
13 T	Acrolein	0.052	0.046	11.5	104	0.00
14 T	Allyl chloride	0.965	0.957	0.8	104	0.00
15 T	Acrylonitrile	0.250	0.256	-2.4	106	0.00
16 T	Acetone	0.205	0.200	2.4	106	0.00
17 T	Carbon Disulfide	1.422	1.332	6.3	101	0.00
18 T	Methyl Acetate	0.541	0.577	-6.7	107	0.00
19 T	Methyl tert-butyl Ether	1.531	1.598	-4.4	107	0.00
20 T	Methylene Chloride	0.607	0.596	1.8	104	0.00
21 T	trans-1,2-Dichloroethene	0.555	0.547	1.4	103	0.00
22 T	Diisopropyl ether	1.865	1.923	-3.1	103	0.00
23 T	Vinyl Acetate	1.397	1.476	-5.7	106	0.00
24 P	1,1-Dichloroethane	1.064	1.052	1.1	103	0.00
25 T	2-Butanone	0.274	0.286	-4.4	106	0.00
26 T	2,2-Dichloropropane	0.792	0.781	1.4	105	0.00
27 T	cis-1,2-Dichloroethene	0.677	0.659	2.7	102	0.00
28 T	Bromochloromethane	0.525	0.527	-0.4	102	0.00
29 T	Tetrahydrofuran	0.175	0.187	-6.9	109	0.00
30 C	Chloroform	1.087	1.051	3.3#	102	0.00
31 T	Cyclohexane	0.944	0.932	1.3	104	0.00
32 T	1,1,1-Trichloroethane	0.908	0.891	1.9	103	0.00
33 S	1,2-Dichloroethane-d4	0.661	0.645	2.4	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.339	0.342	-0.9	102	0.00
36 T	1,1-Dichloropropene	0.458	0.451	1.5	104	0.00
37 T	Ethyl Acetate	0.390	0.376	3.6	106	0.00
38 T	Carbon Tetrachloride	0.484	0.478	1.2	102	0.00
39 T	Methylcyclohexane	0.573	0.581	-1.4	104	0.00
40 TM	Benzene	1.385	1.368	1.2	102	0.00
41 T	Methacrylonitrile	0.208	0.226	-8.7	108	0.00
42 TM	1,2-Dichloroethane	0.470	0.462	1.7	102	0.00
43 T	Isopropyl Acetate	0.597	0.640	-7.2	105	0.00
44 TM	Trichloroethene	0.361	0.348	3.6	103	0.00
45 C	1,2-Dichloropropane	0.350	0.349	0.3#	103	0.00
46 T	Dibromomethane	0.242	0.240	0.8	102	0.00
47 T	Bromodichloromethane	0.518	0.513	1.0	102	0.00
48 T	Methyl methacrylate	0.308	0.339	-10.1	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070225\
 Data File : VX046867.D
 Acq On : 02 Jul 2025 15:14
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX070225

Quant Time: Jul 03 05:55:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	108	0.00
50 S	Toluene-d8	1.197	1.199	-0.2	102	0.00
51 T	4-Methyl-2-Pentanone	0.353	0.381	-7.9	106	0.00
52 CM	Toluene	0.858	0.859	-0.1#	104	0.00
53 T	t-1,3-Dichloropropene	0.462	0.489	-5.8	105	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.549	-4.2	104	0.00
55 T	1,1,2-Trichloroethane	0.320	0.319	0.3	102	0.00
56 T	Ethyl methacrylate	0.442	0.482	-9.0	107	0.00
57 T	1,3-Dichloropropane	0.550	0.547	0.5	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.268	-7.6	102	0.00
59 T	2-Hexanone	0.234	0.260	-11.1	107	0.00
60 T	Dibromochloromethane	0.383	0.383	0.0	103	0.00
61 T	1,2-Dibromoethane	0.321	0.329	-2.5	104	0.00
62 S	4-Bromofluorobenzene	0.455	0.463	-1.8	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.333	0.318	4.5	103	0.00
65 PM	Chlorobenzene	1.081	1.064	1.6	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.363	0.0	103	0.00
67 C	Ethyl Benzene	1.851	1.845	0.3#	103	0.00
68 T	m/p-Xylenes	0.698	0.692	0.9	102	0.00
69 T	o-Xylene	0.674	0.671	0.4	103	0.00
70 T	Styrene	1.153	1.174	-1.8	102	0.00
71 P	Bromoform	0.270	0.268	0.7	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.515	3.567	-1.5	102	0.00
74 T	N-amyl acetate	1.194	1.306	-9.4	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.966	0.978	-1.2	105	0.00
76 T	1,2,3-Trichloropropane	0.773	0.795	-2.8	104	0.00
77 T	Bromobenzene	0.875	0.864	1.3	102	0.00
78 T	n-propylbenzene	4.238	4.258	-0.5	102	0.00
79 T	2-Chlorotoluene	2.503	2.466	1.5	102	0.00
80 T	1,3,5-Trimethylbenzene	2.856	2.937	-2.8	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.301	-4.5	106	0.00
82 T	4-Chlorotoluene	2.888	2.916	-1.0	103	0.00
83 T	tert-Butylbenzene	2.950	3.004	-1.8	103	0.00
84 T	1,2,4-Trimethylbenzene	2.887	2.911	-0.8	102	0.00
85 T	sec-Butylbenzene	3.719	3.751	-0.9	102	0.00
86 T	p-Isopropyltoluene	3.114	3.150	-1.2	103	0.00
87 T	1,3-Dichlorobenzene	1.637	1.623	0.9	103	0.00
88 T	1,4-Dichlorobenzene	1.704	1.630	4.3	103	0.00
89 T	n-Butylbenzene	2.926	2.970	-1.5	103	0.00
90 T	Hexachloroethane	0.552	0.552	0.0	103	0.00
91 T	1,2-Dichlorobenzene	1.587	1.541	2.9	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.169	0.176	-4.1	107	0.00
93 T	1,2,4-Trichlorobenzene	1.078	1.094	-1.5	104	0.00
94 T	Hexachlorobutadiene	0.407	0.428	-5.2	108	0.00
95 T	Naphthalene	2.840	3.069	-8.1	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070225\
Data File : VX046867.D
Acq On : 02 Jul 2025 15:14
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX070225

Quant Time: Jul 03 05:55:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.019	1.036	-1.7	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6