

Surrogate Summary

SDG No.: **Q2600**

Client: **T&A Construction Inc**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2600-04	TRENCH	1,2-Dichloroethane-d4	50	53.4	107		74	125
		Dibromofluoromethane	50	48.0	96		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	55.1	110		77	121
Q2600-08	STOCK-PILE	1,2-Dichloroethane-d4	50	55.0	110		74	125
		Dibromofluoromethane	50	48.9	98		75	124
		Toluene-d8	50	50.7	101		86	113
		4-Bromofluorobenzene	50	55.5	111		77	121
Q2600-12	END-OF-TRENCH	1,2-Dichloroethane-d4	50	51.2	102		74	125
		Dibromofluoromethane	50	47.6	95		75	124
		Toluene-d8	50	49.3	99		86	113
		4-Bromofluorobenzene	50	53.0	106		77	121
VX0716WBL01	VX0716WBL01	1,2-Dichloroethane-d4	50	53.6	107		74	125
		Dibromofluoromethane	50	49.5	99		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	52.5	105		77	121
VX0716WBS01	VX0716WBS01	1,2-Dichloroethane-d4	50	52.4	105		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121
VX0716WBSD01	VX0716WBSD01	1,2-Dichloroethane-d4	50	55.4	111		74	125
		Dibromofluoromethane	50	50.2	100		75	124
		Toluene-d8	50	48.0	96		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121