

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2600</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>T&A Construction Inc</u>	Datafile :	<u>VX047023.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0716WBSD01	Vinyl chloride	20	17.0	ug/L	85	2		65	117	19
	1,1-Dichloroethene	20	17.9	ug/L	90	1		74	110	20
	2-Butanone	100	150	ug/L	150	7	*	65	122	26
	Carbon Tetrachloride	20	18.3	ug/L	92	0		77	113	15
	Chloroform	20	20.7	ug/L	104	4		79	113	20
	Benzene	20	18.5	ug/L	93	0		82	109	15
	1,2-Dichloroethane	20	20.8	ug/L	104	3		80	115	20
	Trichloroethene	20	17.5	ug/L	88	0		77	113	15
	Tetrachloroethene	20	17.6	ug/L	88	0		67	123	15
	Chlorobenzene	20	19.3	ug/L	97	2		82	109	15