

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TACO01
 Lab Code: ACE SDG NO.: Q2600
 Lab File ID: VX047016.D Date Analyzed: 07/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	329194	5.56	557691	6.76	489553	10.06
UPPER LIMIT	658388	6.056	1115380	7.263	979106	10.555
LOWER LIMIT	164597	5.056	278846	6.263	244777	9.555
EPA SAMPLE NO.						
TRENCH	456381	5.56	801649	6.77	755525	10.06
STOCK-PILE	466941	5.57	838214	6.77	804001	10.06
END-OF-TRENCH	367687	5.57	638586	6.77	581092	10.06
VX0716WBL01	324808	5.56	553162	6.77	506524	10.06
VX0716WBS01	322183	5.57	547259	6.77	491570	10.06
VX0716WBSD01	310963	5.56	532754	6.77	475577	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.