

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TACO01
 Lab Code: ACE SDG NO.: Q2600
 Lab File ID: VX046988.D Date Analyzed: 07/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	298555	5.56	499158	6.76	443564	10.06
UPPER LIMIT	597110	6.056	998316	7.263	887128	10.555
LOWER LIMIT	149278	5.056	249579	6.263	221782	9.555
EPA SAMPLE NO.						
TRENCH	342096	5.56	576254	6.76	518427	10.06
STOCK-PILE	340958	5.56	581697	6.76	519348	10.06
END-OF-TRENCH	300812	5.56	484192	6.76	369331	10.06
VX0715MBL01	342671	5.56	579223	6.77	530422	10.06
VX0715MBS01	310444	5.57	528338	6.77	471044	10.06
VX0715MBSD01	285726	5.57	492414	6.77	437242	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.