

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TACO01
 Lab Code: ACE SDG NO.: Q2600
 Lab File ID: VX047016.D Date Analyzed: 07/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	329194	5.56	557691	6.76	489553	10.06
UPPER LIMIT	658388	6.056	1115380	7.263	979106	10.555
LOWER LIMIT	164597	5.056	278846	6.263	244777	9.555
EPA SAMPLE NO.						
END-OF-TRENCHDL	320742	5.56	537318	6.77	482473	10.06
VX0716MBL01	319061	5.56	562969	6.77	523286	10.06
VX0716MBS01	328600	5.56	550527	6.76	493708	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.