

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TACO01
 Lab Code: ACE SDG NO.: Q2600
 Lab File ID: VX046988.D Date Analyzed: 07/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	217822	12.018				
UPPER LIMIT	435644	12.518				
LOWER LIMIT	108911	11.518				
EPA SAMPLE NO.						
TRENCH	226022	12.02				
STOCK-PILE	233909	12.02				
END-OF-TRENCH	180753	12.02				
VX0715MBL01	275487	12.02				
VX0715MBS01	239011	12.02				
VX0715MBSD01	224565	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.