

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2603</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>07/16/2025 15:34</u>
Lab File ID:	<u>BN037518.D</u>	Init. Calib. Date(s):	<u>07/15/2025 07/15/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>12:36 16:14</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.525		-8.5	50.0
Fluoranthene-d10	1.060	0.943		-11.0	50.0
2-Fluorophenol	0.989	0.896		-9.4	50.0
Phenol-d6	1.241	1.092		-12.0	50.0
Nitrobenzene-d5	0.299	0.280		-6.4	50.0
2-Fluorobiphenyl	2.080	2.170		4.3	50.0
2,4,6-Tribromophenol	0.197	0.155		-21.3	50.0
Terphenyl-d14	0.859	0.847		-1.4	50.0
1,4-Dioxane	0.385	0.391		1.6	50.0

All other compounds must meet a minimum RRF of 0.010.