

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037509.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2603
DFTPP Injection Date: 07/16/2025
DFTPP Injection Time: 09:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	36.7
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	14.6
442	Greater than 50% of mass 198	98
443	15.0 - 24.0% of mass 442	17.5 (17.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037510.D	07/16/2025	09:50
PB168868BL	PB168868BL	BN037511.D	07/16/2025	10:27
RW7-SP303-Test-20250709	Q2603-01	BN037512.D	07/16/2025	11:04
RW7-Pond-Test-20250709	Q2603-02	BN037513.D	07/16/2025	11:40
PB168868BS	PB168868BS	BN037516.D	07/16/2025	13:29
PB168868BSD	PB168868BSD	BN037517.D	07/16/2025	14:06
SSTDCCC0.4EC	SSTDCCC0.4	BN037518.D	07/16/2025	15:34