

## **DATA PACKAGE GENERAL CHEMISTRY**

**PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13**

**TETRA TECH NUS, INC.**

**661 Andersen Drive**

**Suite 200**

**Pittsburgh, PA - 15220-2745**

**Phone No: 412-921-7090**

**ORDER ID : Q2604**

**ATTENTION : Ernie Wu**



**Laboratory Certification ID # 20012**



1) GENERAL CHEMISTRY DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	5
5) Conformance/Non Conformance	6
6) QA Checklist	7
7) Chronicle	8
8) Sample Data	9
8.1) RW8-SP303-20250711	10
9) QC Data Summary For Genchem	11
9.1) Preparation Blank Summary	12
9.2) Duplicate Sample Summary	13
9.3) Laboratory Control Sample Summary	15
10) GENCHEM RAW DATA	17
10.1) GENCHEM RAW DATA - ANALYTICAL	18
10.1.1) LB136528	18
10.1.2) LB136530	20
11) Analytical Runlogs	22
12) Standard Prep Logs	24
13) Shipping Document	26
13.1) Chain Of Custody	27
13.2) Lab Certificate	28

1
2
3
4
5
6
7
8
9
10
11
12
13

## Cover Page

**Order ID :** Q2604

**Project ID :** NWIRP Bethpage 112G08005-WE13

**Client :** Tetra Tech NUS, Inc.

**Lab Sample Number**

Q2604-01  
Q2604-02

**Client Sample Number**

RW8-SP100-20250711  
RW8-SP303-20250711

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 1:45 pm, Jul 25, 2025*

Date: 7/18/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Tetra Tech NUS, Inc.**

**Project Name: NWIRP Bethpage 112G08005-WE13**

**Project Manager : Ernie Wu**

**Order ID # Q2604**

**Test Name: TDS,TSS**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 07/15/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: TDS,TSS. This data package contains results for TDS,TSS.

### **C. Analytical Techniques:**

The analysis of TDS was based on method SM2540 C and The analysis of TSS was based on method SM2540 D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 1:45 pm, Jul 25, 2025*

Signature \_\_\_\_\_

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q2604

MATRIX: Water

METHOD: SM2540 C,SM2540 D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. The Blank Spike met requirements for all compounds.			✓
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

QA REVIEW

**REVIEWED**

*By Sohil Jodhani, QA/QC Director at 10:23 am, Jul 25, 2025*

APPENDIX A

**QA REVIEW GENERAL DOCUMENTATION**

Project #: Q2604

Completed

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

**ANALYTICAL:**

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/18/2025

## LAB CHRONICLE

<b>OrderID:</b>	Q2604	<b>OrderDate:</b>	7/15/2025 10:29:00 AM
<b>Client:</b>	Tetra Tech NUS, Inc.	<b>Project:</b>	NWIRP Bethpage 112G08005-WE13
<b>Contact:</b>	Ernie Wu	<b>Location:</b>	O11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2604-02	RW8-SP303-2025071 1	WATER			07/11/25 13:13			07/15/25
			TDS	SM2540 C			07/17/25 12:30	
			TSS	SM2540 D			07/17/25 15:00	





# SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13

## Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/11/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/15/25
Client Sample ID:	RW8-SP303-20250711	SDG No.:	Q2604
Lab Sample ID:	Q2604-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	11.0		1	1.00	10.0	10.0	mg/L		07/17/25 12:30	SM 2540 C-20
TSS	4.40		1	1.00	4.00	4.00	mg/L		07/17/25 15:00	SM 2540 D-20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



# QC RESULT SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

## Preparation Blank Summary

**Client:** Tetra Tech NUS, Inc.

**SDG No.:** Q2604

**Project:** NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>LB136528BL</b>							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	07/17/2025
Sample ID: <b>LB136530BL</b>							
TSS	mg/L	1	2.0000	J	1	4	07/17/2025

### Duplicate Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2604
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Sample ID:</b>	Q2604-02
<b>Client ID:</b>	RW8-SP303-20250711DUP	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	11.0		11.0		1	0		07/17/2025

### Duplicate Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2604
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Sample ID:</b>	Q2615-02
<b>Client ID:</b>	COMPDUP	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	145		144		1	0.69		07/17/2025

### Laboratory Control Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2604
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Run No.:</b>	LB136528

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136528BS							
TDS	mg/L	100	95.0		95	1	90-110	07/17/2025

### Laboratory Control Sample Summary

<b>Client:</b>	Tetra Tech NUS, Inc.	<b>SDG No.:</b>	Q2604
<b>Project:</b>	NWIRP Bethpage 112G08005-WE13	<b>Run No.:</b>	LB136530

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136530BS							
TSS	mg/L	550	533		97	1	90-110	07/17/2025





# RAW DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

# TOTAL Dissolved Solids - SM2540C

Run Number: LB136528

Date: 07/18/2025

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

Filter ID: 17416528

EMPTY DISH				EMPTY DISH				OvenID:	WC OVEN-1	Thermo ID:	WET OVEN#1
TEMP IN:	104 °C	07/17/2025	11:00	TEMP OUT:	104 °C	07/17/2025	12:00				
TEMP1 IN:	104 °C	07/17/2025	12:30	TEMP1 OUT:	104 °C	07/17/2025	13:30	OvenID1:	WC OVEN-1	Thermo ID1:	WET OVEN#1
TEMP2 IN:	104 °C	07/17/2025	16:00	TEMP2 OUT:	103 °C	07/18/2025	07:30	OvenID2:	WC OVEN-2	Thermo ID2:	WET OVEN#2
TEMP3 IN:	180 °C	07/18/2025	07:35	TEMP3 OUT:	182 °C	07/18/2025	09:00	OvenID3:	WC OVEN-2	Thermo ID3:	WET OVEN#2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180(±2) °C (g)	Weight (g)	Result mg/L
1	LB136528BL	LB136528BL	65.7812	65.7812	100	65.7812	65.7812	65.7812	0.0000	0
2	LB136528BS	LB136528BS	85.3036	85.3036	100	85.3131	85.3131	85.3131	0.0095	95
3	Q2604-02	RW8-SP303-20250711	106.1519	106.1520	100	106.1531	106.1531	106.1531	0.0011	11
4	Q2604-02DUP	RW8-SP303-20250711DUP	110.2244	110.2244	100	110.2255	110.2255	110.2255	0.0011	11

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180(±2) °C(g)

D = Weight (g)

Q2604-GENCHEM

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WB136528

WorkList Name : tds q2604

WorkList ID : 190798

Department : Wet-Chemistry

Date : 07-17-2025 10:28:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2604-02	RW8-SP303-20250711	Water	TDS	Cool 4 deg C	TETR06	O11	07/11/2025	SM2540 C

Date/Time 07-17-25 12:00

Raw Sample Received by: JB WCI

Raw Sample Relinquished by: JTCsm

Date/Time 07-17-25 16:30

Raw Sample Received by: JTCsm

Raw Sample Relinquished by: JB WCI

# TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 07/17/2025

Run Number: LB136530

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 07/17/2025 10:00 TEMP1 OUT: 103 °C 07/17/2025 11:00  
 TEMP2 IN: 104 °C 07/17/2025 12:00 TEMP2 OUT: 104 °C 07/17/2025 12:35  
 TEMP3 IN: 104 °C 07/17/2025 15:00 TEMP3 OUT: 103 °C 07/17/2025 16:30  
 TEMP4 IN: 104 °C 07/17/2025 17:00 TEMP4 OUT: 103 °C 07/17/2025 18:35

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB136530BL	LB136530BL	1.3562	1.3562	100	1.3563	1.3563	1.3563	0.0001	1
2	LB136530BS	LB136530BS	1.5843	1.5843	100	1.6376	1.6376	1.6376	0.0533	533
3	Q2604-02	RW8-SP303-20250711	1.4849	1.4849	250	1.4860	1.4860	1.4860	0.0011	4.4
4	Q2615-02	COMP	1.4654	1.4654	200	1.4944	1.4944	1.4944	0.0290	145
5	Q2615-02DUP	COMPDUP	1.4754	1.4754	200	1.5042	1.5042	1.5042	0.0288	144
6	Q2629-01	011 willets Pt Blvd(may)	1.4888	1.4888	150	1.5199	1.5199	1.5199	0.0311	207.3
7	Q2629-02	022 35th Ave(may)	1.4954	1.4954	150	1.5240	1.5240	1.5240	0.0286	190.7
8	Q2630-01	001 Willets Pt Blvd (June)	1.4692	1.4692	200	1.4937	1.4937	1.4937	0.0245	122.5
9	Q2630-02	002 35th Ave (JUNE)	1.5002	1.5002	200	1.5252	1.5252	1.5252	0.0250	125
10	Q2633-02	FIBER-GLASS-TANK	1.4761	1.4761	2000	1.4765	1.4765	1.4765	0.0004	0.2

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L =  $\frac{D}{A} \times 1000 \times 1000$

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss q2629

WorkList ID : 190799

Department : Wet-Chemistry

Date : 07-17-2025 11:17:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2604-02 <u>B</u>	RW8-SP303-20250711	Water	TSS	Cool 4 deg C	TETR06	O11	07/11/2025	SM2540 D
Q2615-02 <u>A</u>	COMP	Water	TSS	Cool 4 deg C	ARAM01	O11	07/16/2025	SM2540 D
Q2629-01 <u>B</u>	011 willets Pt Blvd(may)	Water	TSS	Cool 4 deg C	TULL01	O32	07/16/2025	SM2540 D
Q2629-02 <u>B</u>	022 35th Ave(may)	Water	TSS	Cool 4 deg C	TULL01	O32	07/16/2025	SM2540 D
Q2630-01 <u>D</u>	001 Willets Pt Blvd (June)	Water	TSS	Cool 4 deg C	TULL01	O11	07/16/2025	SM2540 D
Q2630-02 <u>D</u>	002 35th Ave (JUNE)	Water	TSS	Cool 4 deg C	TULL01	O11	07/16/2025	SM2540 D
Q2633-02 <u>124</u>	FIBER-GLASS-TANK	Water	TSS	Cool 4 deg C	PSEG03	D41	07/17/2025	SM2540 D

UB 136530

Date/Time 07/17/25 13:10  
 Raw Sample Received by: JO WDC  
 Raw Sample Relinquished by: ITCSM

Date/Time 07/17/25Raw Sample Received by: ITCSMRaw Sample Relinquished by: ITCSM

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136528**

Review By	jignesh	Review On	7/18/2025 8:28:14 AM
Supervise By	Iwona	Supervise On	7/18/2025 11:01:53 AM
SubDirectory	LB136528	Test	TDS
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136528BL	LB136528BL	MB	07/17/25 12:30		jignesh	OK
2	LB136528BS	LB136528BS	LCS	07/17/25 12:30	WP113986	jignesh	OK
3	Q2604-02	RW8-SP303-2025071	SAM	07/17/25 12:30		jignesh	OK
4	Q2604-02DUP	RW8-SP303-2025071	DUP	07/17/25 12:30		jignesh	OK

Instrument ID: WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136530**

Review By	jignesh	Review On	7/18/2025 10:53:55 AM
Supervise By	Iwona	Supervise On	7/18/2025 11:01:25 AM
SubDirectory	LB136530	Test	TSS
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136530BL	LB136530BL	MB	07/17/25 15:00		jignesh	OK
2	LB136530BS	LB136530BS	LCS	07/17/25 15:00	55MG W3186 + 100 ML W3112	jignesh	OK
3	Q2604-02	RW8-SP303-2025071	SAM	07/17/25 15:00		jignesh	OK
4	Q2615-02	COMP	SAM	07/17/25 15:00		jignesh	OK
5	Q2615-02DUP	COMPDUP	DUP	07/17/25 15:00		jignesh	OK
6	Q2629-01	011 willets Pt Blvd(ma	SAM	07/17/25 15:00		jignesh	OK
7	Q2629-02	022 35th Ave(may)	SAM	07/17/25 15:00		jignesh	OK
8	Q2630-01	001 Willets Pt Blvd (Ju	SAM	07/17/25 15:00		jignesh	OK
9	Q2630-02	002 35th Ave (JUNE)	SAM	07/17/25 15:00		jignesh	OK
10	Q2633-02	FIBER-GLASS-TANK	SAM	07/17/25 15:00		jignesh	OK

### Prep Standard - Chemical Standard Summary

**Order ID :** Q2604

**Test :** TDS,TSS

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** LB136528, LB136530,

**Standard ID :**

**Chemical ID :**

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13



### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
				/	/	

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13



# SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

<b>CHEMTECH</b> CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: <span style="font-size: 1.2em;">Q 2604</span>																					
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION																					
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: PO#																					
ADDRESS: 4433 Corporation Ln, Suite 300		PROJECT #: 112G08005-WE13 LOCATION: RW8		ADDRESS:																					
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:																					
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetratech.com		ATTENTION: PHONE:																					
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148																							
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS																					
FAX: _____ 10 _____ DAYS* HARD COPY: _____ 10 _____ DAYS* EDD _____ 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		<input type="checkbox"/> RESEULTS ONLY <input type="checkbox"/> USEPA CLP <input type="checkbox"/> RESULTS + QC <input type="checkbox"/> New York State ASP "B" <input type="checkbox"/> New Jersey REDUCED <input type="checkbox"/> New York State ASP "A" <input type="checkbox"/> New Jersey CLP <input type="checkbox"/> Other _____ <input type="checkbox"/> EDD Format _____		<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">1,4-Dioxane SW846 8270</td> <td style="width: 10%;">Iron, Total</td> <td style="width: 10%;">TSS</td> <td style="width: 10%;">TDS</td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> </tr> <tr> <td>1</td> <td>2</td> <td>3</td> <td>4</td> <td>5</td> <td>6</td> <td>7</td> <td>8</td> <td>9</td> <td></td> </tr> </table>		1,4-Dioxane SW846 8270	Iron, Total	TSS	TDS							1	2	3	4	5	6	7	8	9	
1,4-Dioxane SW846 8270	Iron, Total	TSS	TDS																						
1	2	3	4	5	6	7	8	9																	
				PRESERVATIVES																					
				COMMENTS																					
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP GRAB	SAMPLE COLLECTION DATE TIME	# of Bottles																				
1.	RW8-SP100-20250711	GW		7/11/25 13:05	2																				
2.	RW8-SP303-20250711	GW		7/11/25 13:13	4																				
3.																									
4.																									
5.																									
6.																									
7.																									
8.																									
9.																									
10.																									

**SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY**

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: <input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant <input type="checkbox"/> Cooler Temp <u>24.5</u> MeOH extraction requires an additional 4oz. Jar for percent solid <input type="checkbox"/> Ice in Cooler? <u>yes</u>
1.	7/14/25 13:00	1.	Comments:  <div style="display: flex; justify-content: space-between; align-items: flex-end;"> <div>           SHIPPED VIA: CLIENT: <input type="checkbox"/> Hand Delivered    <input type="checkbox"/> Overnight            CHEMTECH:    <input type="checkbox"/> Picked Up    <input type="checkbox"/> Overnight         </div> <div> <b>Shipment Complete</b>  <input type="checkbox"/> YES    <input type="checkbox"/> NO         </div> </div>
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.	7/15/25 10:15	2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	Page _____ of _____
3.		3.	

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT    YELLOW - CHEMTECH COPY    PINK - SAMPLER COPY

### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488