

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2604

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4310	4000	108	90 - 110	P	07/17/2025	13:30	LB136533

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2604

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	110	100	110	80 - 120	P	07/17/2025	13:37	LB136533

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2604

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5040	5000	101	90 - 110	P	07/17/2025	14:20	LB136533
CCV02	Iron	5080	5000	102	90 - 110	P	07/17/2025	14:58	LB136533
CCV03	Iron	5030	5000	101	90 - 110	P	07/17/2025	15:50	LB136533
CCV04	Iron	5040	5000	101	90 - 110	P	07/17/2025	16:42	LB136533
CCV05	Iron	4970	5000	99	90 - 110	P	07/17/2025	17:58	LB136533
CCV06	Iron	4770	5000	95	90 - 110	P	07/17/2025	19:24	LB136533
CCV07	Iron	4890	5000	98	90 - 110	P	07/17/2025	20:16	LB136533