

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q2604

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	98300	101000	97	85600	116500	07/17/2025	13:51	LB136533
ICSAB01	Iron	97600	99300	98	84400	114500	07/17/2025	13:56	LB136533