

LAB CHRONICLE

OrderID:	Q2604	OrderDate:	7/15/2025 10:29:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	O11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2604-01	RW8-SP100-2025071 1	Water			07/11/25			07/15/25
			SVOC-SIMGroup1	8270-Modified		07/15/25	07/16/25	
Q2604-02	RW8-SP303-2025071 1	Water			07/11/25			07/15/25
			SVOC-SIMGroup1	8270-Modified		07/15/25	07/16/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2604
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
Total Svoc :					0.00				
Total Concentration:					0.00				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/15/25
Client Sample ID:	RW8-SP100-20250711	SDG No.:	Q2604
Lab Sample ID:	Q2604-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037514.D	1	07/15/25 11:40	07/16/25 12:16	PB168868

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		90%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		101%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.44		58 - 132		109%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1760	7.724				
1146-65-2	Naphthalene-d8	4260	10.509				
15067-26-2	Acenaphthene-d10	2260	14.355				
1517-22-2	Phenanthrene-d10	4520	17.099				
1719-03-5	Chrysene-d12	3750	21.277				
1520-96-3	Perylene-d12	3700	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/15/25
Client Sample ID:	RW8-SP303-20250711	SDG No.:	Q2604
Lab Sample ID:	Q2604-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037515.D	1	07/15/25 11:40	07/16/25 12:53	PB168868

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		95%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.33		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.42		53 - 106		105%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53		58 - 132		131%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1670	7.725				
1146-65-2	Naphthalene-d8	4000	10.509				
15067-26-2	Acenaphthene-d10	2070	14.356				
1517-22-2	Phenanthrene-d10	4340	17.099				
1719-03-5	Chrysene-d12	3980	21.277				
1520-96-3	Perylene-d12	3910	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2604

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168868BL	PB168868BL	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.33	81		30	150
		Nitrobenzene-d5	0.4	0.37	91		55	111
		2-Fluorobiphenyl	0.4	0.38	95		53	106
		Terphenyl-d14	0.4	0.43	107		58	132
PB168868BS	PB168868BS	2-Methylnaphthalene-d10	0.4	0.37	93		30	150
		Fluoranthene-d10	0.4	0.31	79		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.40	99		58	132
PB168868BSD	PB168868BSD	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
Q2604-01	RW8-SP100-20250711	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.40	101		53	106
		Terphenyl-d14	0.4	0.44	109		58	132
Q2604-02	RW8-SP303-20250711	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.33	81		55	111
		2-Fluorobiphenyl	0.4	0.42	105		53	106
		Terphenyl-d14	0.4	0.53	131		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2604

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037516.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168868BS	1,4-Dioxane	0.4	0.31	ug/L	78				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2604

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037517.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168868BSD	1,4-Dioxane	0.4	0.29	ug/L	73	7			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB168868BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2604

Lab File ID: BN037511.D

Lab Sample ID: PB168868BL

Instrument ID: BNA_N

Date Extracted: 07/15/2025

Matrix: (soil/water) Water

Date Analyzed: 07/16/2025

Level: (low/med) LOW

Time Analyzed: 10:27

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168868BS	PB168868BS	BN037516.D	07/16/2025
RW8-SP100-20250711	Q2604-01	BN037514.D	07/16/2025
RW8-SP303-20250711	Q2604-02	BN037515.D	07/16/2025
PB168868BSD	PB168868BSD	BN037517.D	07/16/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037497.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2604
DFTPP Injection Date: 07/15/2025
DFTPP Injection Time: 10:57

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	83.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037499.D	07/15/2025	12:36
SSTDICC0.2	SSTDICC0.2	BN037500.D	07/15/2025	13:12
SSTDICCC0.4	SSTDICCC0.4	BN037501.D	07/15/2025	13:49
SSTDICC0.8	SSTDICC0.8	BN037502.D	07/15/2025	14:25
SSTDICC1.6	SSTDICC1.6	BN037503.D	07/15/2025	15:01
SSTDICC3.2	SSTDICC3.2	BN037504.D	07/15/2025	15:38
SSTDICC5.0	SSTDICC5.0	BN037505.D	07/15/2025	16:14

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037509.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2604
DFTPP Injection Date: 07/16/2025
DFTPP Injection Time: 09:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	83.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.5 (17.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037510.D	07/16/2025	09:50
PB168868BL	PB168868BL	BN037511.D	07/16/2025	10:27
RW8-SP100-20250711	Q2604-01	BN037514.D	07/16/2025	12:16
RW8-SP303-20250711	Q2604-02	BN037515.D	07/16/2025	12:53
PB168868BS	PB168868BS	BN037516.D	07/16/2025	13:29
PB168868BSD	PB168868BSD	BN037517.D	07/16/2025	14:06
SSTDCCC0.4EC	SSTDCCC0.4	BN037518.D	07/16/2025	15:34

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2604

Client ID : SSTDCCC0.4

Date Analyzed: 07/16/2025

Lab File ID: BN037510.D

Time Analyzed: 09:50

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2448	7.724	6343	10.51	3548	14.36
UPPER LIMIT	4896	8.224	12686	11.009	7096	14.855
LOWER LIMIT	1224	7.224	3171.5	10.009	1774	13.855
EPA SAMPLE NO.						
01 PB168868BL	3101	7.72	7573	10.51	3896	14.36
02 PB168868BS	1941	7.72	4724	10.51	2338	14.36
03 PB168868BSD	1974	7.72	4747	10.51	2348	14.36
04 RW8-SP100-20250711	1757	7.72	4258	10.51	2264	14.36
05 RW8-SP303-20250711	1670	7.73	3997	10.51	2071	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2604

Client ID: SSTDCCC0.4

Date Analyzed: 07/16/2025

Lab File ID: BN037510.D

Time Analyzed: 09:50

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6984	17.086	5489	21.277	5154	23.513
UPPER LIMIT	13968	17.586	10978	21.777	10308	24.013
LOWER LIMIT	3492	16.586	2744.5	20.777	2577	23.013
EPA SAMPLE NO.						
01 PB168868BL	7545	17.10	5013	21.28	4525	23.51
02 PB168868BS	4406	17.10	3170	21.28	2673	23.52
03 PB168868BSD	4399	17.10	3154	21.28	2636	23.52
04 RW8-SP100-20250711	4524	17.10	3752	21.28	3699	23.52
05 RW8-SP303-20250711	4336	17.10	3975	21.28	3913	23.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168868BL		SDG No.:	Q2604
Lab Sample ID:	PB168868BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037511.D	1	07/15/25 11:40	07/16/25 10:27	PB168868

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		84%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		81%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		91%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		95%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		107%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	3100	7.724				
1146-65-2	Naphthalene-d8	7570	10.509				
15067-26-2	Acenaphthene-d10	3900	14.355				
1517-22-2	Phenanthrene-d10	7550	17.099				
1719-03-5	Chrysene-d12	5010	21.277				
1520-96-3	Perylene-d12	4530	23.513				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168868BS		SDG No.:	Q2604
Lab Sample ID:	PB168868BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037516.D	1	07/15/25 11:40	07/16/25 13:29	PB168868

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.31		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.37		30 - 150		93%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.31		30 - 150		79%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		89%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		98%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		99%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1940	7.724				
1146-65-2	Naphthalene-d8	4720	10.509				
15067-26-2	Acenaphthene-d10	2340	14.355				
1517-22-2	Phenanthrene-d10	4410	17.099				
1719-03-5	Chrysene-d12	3170	21.277				
1520-96-3	Perylene-d12	2670	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168868BSD		SDG No.:	Q2604
Lab Sample ID:	PB168868BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037517.D	1	07/15/25 11:40	07/16/25 14:06	PB168868

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.29		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		89%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		76%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		98%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		95%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1970	7.724				
1146-65-2	Naphthalene-d8	4750	10.509				
15067-26-2	Acenaphthene-d10	2350	14.356				
1517-22-2	Phenanthrene-d10	4400	17.099				
1719-03-5	Chrysene-d12	3150	21.277				
1520-96-3	Perylene-d12	2640	23.519				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN071525.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 16 02:38:11 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037499.D 0.2 =BN037500.D 0.4 =BN037501.D 0.8 =BN037502.D 1.6 =BN037503.D 3.2 =BN037504.D 5 =BN037505.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.409	0.395	0.371	0.398	0.380	0.354	0.385		5.29
3)	n-Nitrosodimet...	0.466	0.464	0.465	0.508	0.499	0.501	0.484		4.31
4) S	2-Fluorophenol	1.038	1.011	0.985	0.908	0.982	0.971	1.030	0.989	4.42
5) S	Phenol-d6	1.448	1.238	1.190	1.105	1.201	1.229	1.275	1.241	8.52
6)	bis(2-Chloroet...	1.082	1.052	1.024	0.983	1.037	1.033	1.016	1.033	2.99
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.311	0.288	0.283	0.270	0.300	0.305	0.336	0.299	7.20
9)	Naphthalene	1.069	1.054	1.046	1.009	1.091	1.073	1.126	1.067	3.45
10)	Hexachlorobuta...	0.229	0.237	0.235	0.223	0.245	0.236	0.246	0.236	3.44
11) SURR	2-Methylnaphth...	0.556	0.534	0.541	0.522	0.562	0.590	0.711	0.574	11.24
12)	2-Methylnaphth...	0.704	0.655	0.678	0.665	0.716	0.736	0.756	0.701	5.34
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.197	0.173	0.173	0.176	0.194	0.215	0.248	0.197	13.98
15) S	2-Fluorobiphenyl	1.818	1.794	2.045	2.024	2.277	2.205	2.397	2.080	10.91
16)	Acenaphthylene	1.723	1.708	1.719	1.684	1.830	1.895	1.981	1.792	6.30
17)	Acenaphthene	1.239	1.160	1.172	1.150	1.238	1.251	1.320	1.218	5.03
18)	Fluorene	1.592	1.488	1.485	1.486	1.605	1.606	1.717	1.569	5.56
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.044	0.041	0.047	0.057	0.070	0.080	0.057		27.89
21)	4-Bromophenyl-...	0.248	0.247	0.243	0.242	0.268	0.272	0.274	0.256	5.58
22)	Hexachlorobenzene	0.315	0.330	0.328	0.321	0.345	0.340	0.338	0.331	3.26
23)	Atrazine	0.173	0.161	0.159	0.158	0.181	0.200	0.220	0.179	13.24
24)	Pentachlorophenol	0.131	0.125	0.126	0.151	0.170	0.189	0.149		17.64
25)	Phenanthrene	1.167	1.163	1.160	1.129	1.248	1.248	1.273	1.198	4.70
26)	Anthracene	1.025	1.025	1.013	1.023	1.160	1.176	1.232	1.093	8.45
27) SURR	Fluoranthene-d10	1.023	0.998	0.962	0.928	1.041	1.078	1.385	1.060	14.34
28)	Fluoranthene	1.358	1.310	1.290	1.270	1.429	1.431	1.585	1.382	7.96
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.754	1.559	1.607	1.549	1.607	1.665	1.539	1.612	4.74
31) S	Terphenyl-d14	0.926	0.815	0.844	0.811	0.854	0.902	0.865	0.859	4.94
32)	Benzo(a)anthra...	1.414	1.357	1.341	1.285	1.429	1.464	1.517	1.401	5.63
33)	Chrysene	1.452	1.461	1.434	1.358	1.488	1.490	1.528	1.459	3.70
34)	Bis(2-ethylhex...	0.603	0.564	0.538	0.603	0.693	0.779	0.630		14.26
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN071525.M

36)	Indeno(1,2,3-c...	1.493	1.528	1.514	1.559	1.771	1.805	1.991	1.666	11.48
37)	Benzo(b)fluora...	1.464	1.378	1.454	1.436	1.589	1.617	1.692	1.518	7.53
38)	Benzo(k)fluora...	1.516	1.420	1.486	1.470	1.661	1.689	1.724	1.567	7.75
39) C	Benzo(a)pyrene	1.189	1.152	1.192	1.176	1.320	1.369	1.469	1.267	9.51
40)	Dibenzo(a,h)an...	1.201	1.218	1.216	1.256	1.444	1.483	1.627	1.349	12.46
41)	Benzo(g,h,i)pe...	1.247	1.283	1.309	1.297	1.482	1.497	1.663	1.397	10.98

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2604
 Instrument ID: BNA_N Calibration Date/Time: 07/16/2025 09:50
 Lab File ID: BN037510.D Init. Calib. Date(s): 07/15/2025 07/15/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:36 16:14
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.534		-7.0	20.0
Fluoranthene-d10	1.060	0.921		-13.1	20.0
2-Fluorophenol	0.989	0.945		-4.4	20.0
Phenol-d6	1.241	1.176		-5.2	20.0
Nitrobenzene-d5	0.299	0.281		-6.0	20.0
2-Fluorobiphenyl	2.080	2.104		1.2	20.0
2,4,6-Tribromophenol	0.197	0.158		-19.8	20.0
Terphenyl-d14	0.859	0.818		-4.8	20.0
1,4-Dioxane	0.385	0.402		4.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2604
 Instrument ID: BNA_N Calibration Date/Time: 07/16/2025 15:34
 Lab File ID: BN037518.D Init. Calib. Date(s): 07/15/2025 07/15/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:36 16:14
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.525		-8.5	50.0
Fluoranthene-d10	1.060	0.943		-11.0	50.0
2-Fluorophenol	0.989	0.896		-9.4	50.0
Phenol-d6	1.241	1.092		-12.0	50.0
Nitrobenzene-d5	0.299	0.280		-6.4	50.0
2-Fluorobiphenyl	2.080	2.170		4.3	50.0
2,4,6-Tribromophenol	0.197	0.155		-21.3	50.0
Terphenyl-d14	0.859	0.847		-1.4	50.0
1,4-Dioxane	0.385	0.391		1.6	50.0

All other compounds must meet a minimum RRF of 0.010.