

Surrogate Summary

SW-846

SDG No.: Q2604

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168868BL	PB168868BL	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.33	81		30	150
		Nitrobenzene-d5	0.4	0.37	91		55	111
		2-Fluorobiphenyl	0.4	0.38	95		53	106
		Terphenyl-d14	0.4	0.43	107		58	132
PB168868BS	PB168868BS	2-Methylnaphthalene-d10	0.4	0.37	93		30	150
		Fluoranthene-d10	0.4	0.31	79		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.40	99		58	132
PB168868BSD	PB168868BSD	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
Q2604-01	RW8-SP100-20250711	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.40	101		53	106
		Terphenyl-d14	0.4	0.44	109		58	132
Q2604-02	RW8-SP303-20250711	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.33	81		55	111
		2-Fluorobiphenyl	0.4	0.42	105		53	106
		Terphenyl-d14	0.4	0.53	131		58	132