



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q2606

**Contract:** PSEG04

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| ICB35     | Mercury | 0.027          | +/-0.2              | U            | 0.20 | CV | 07/21/2025       | 14:59            | LB136559      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q2606

**Contract:** PSEG04

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB32     | Mercury | 0.027          | +/-0.2              | U            | 0.20 | CV | 07/21/2025       | 15:04            | LB136559      |
| CCB33     | Mercury | 0.027          | +/-0.2              | U            | 0.20 | CV | 07/21/2025       | 15:37            | LB136559      |
| CCB34     | Mercury | 0.027          | +/-0.2              | U            | 0.20 | CV | 07/21/2025       | 15:52            | LB136559      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q2606

**Contract:** PSEG04

**Lab Code:** ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 13:41            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 13:41            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 13:41            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 13:41            | LB136533      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q2606

**Contract:** PSEG04

**Lab Code:** ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 14:25            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 14:25            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 14:25            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 14:25            | LB136533      |
| CCB02     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 15:02            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:02            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:02            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:02            | LB136533      |
| CCB03     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 15:56            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:56            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:56            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 15:56            | LB136533      |
| CCB04     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 16:46            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 16:46            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 16:46            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 16:46            | LB136533      |
| CCB05     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 18:02            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 18:02            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 18:02            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 18:02            | LB136533      |
| CCB06     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 19:28            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 19:28            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 19:28            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 19:28            | LB136533      |
| CCB07     | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/17/2025       | 20:20            | LB136533      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/17/2025       | 20:20            | LB136533      |
|           | Thallium | 4.38           | +/-20               | U            | 40.0 | P | 07/17/2025       | 20:20            | LB136533      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/17/2025       | 20:20            | LB136533      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q2606

**Contract:** PSEG04

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|