

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC

SDG No.: Q2606

Contract: PSEG04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	6.52	2.0	326	-18	22	07/17/2025	13:51	LB136533
	Nickel	3.05	2.0	152	-38	42	07/17/2025	13:51	LB136533
	Thallium	9.59			-40	40	07/17/2025	13:51	LB136533
	Zinc	1.87			-40	40	07/17/2025	13:51	LB136533
ICSAB01	Copper	493	511	96	434	588	07/17/2025	13:56	LB136533
	Nickel	988	954	104	810	1100	07/17/2025	13:56	LB136533
	Thallium	103	108	95	68	148	07/17/2025	13:56	LB136533
	Zinc	1060	952	111	809	1095	07/17/2025	13:56	LB136533
ICSA	Copper	5.19	2.0	260	-18	22	07/17/2025	14:01	LB136533
	Nickel	10.8	2.0	542	-38	42	07/17/2025	14:01	LB136533
	Thallium	-5.72			-40	40	07/17/2025	14:01	LB136533
	Zinc	7.25			-40	40	07/17/2025	14:01	LB136533
ICSAB	Copper	519	511	102	434	588	07/17/2025	14:05	LB136533
	Nickel	973	954	102	810	1100	07/17/2025	14:05	LB136533
	Thallium	122	108	113	68	148	07/17/2025	14:05	LB136533
	Zinc	1050	952	110	809	1095	07/17/2025	14:05	LB136533