

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : 151 ALBANY AVE, FREEPORT NY

GFE LLC

58 Nokomis Ave

Lake Hiawatha, NJ - 07034

Phone No: 646-542-3465

ORDER ID : Q2617

ATTENTION : Frank Galdun



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group2- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) VOCMS Group2 Data	8
6) Shipping Document	16
6.1) CHAIN OF CUSTODY	17
6.2) Lab Certificate	21
6.3) Internal COC	22

1
2
3
4
5
6

Cover Page

Order ID : Q2617

Project ID : 151 Albany Ave, Freeport NY

Client : GFE LLC

Lab Sample Number

Q2617-01
Q2617-02
Q2617-03

Client Sample Number

SV1
SV2
SV3

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:50 am, Jul 28, 2025

Date: 7/28/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

GFE LLC

Project Name: 151 Albany Ave, Freeport NY

Project # N/A

Order ID # Q2617

Test Name: VOCMS Group2

A. Number of Samples and Date of Receipt:

3 Air samples were received on 07/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group2. This data package contains results for VOCMS Group2.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax, Entech 7100A Preconcentrator. The analysis of VOCMS Group2 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike for {VL0722ABS01} with File ID: VL042733.D met requirements for all compounds except for Naphthalene[136%] is failing high but no positive hit in associate sample therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements except for Naphthalene have more than 30% RSD in the Initial Calibration with dated 07/22/2025 with L Instrument but as per method two compounds as allowed to be fail less than 40% RSD.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples SV1, SV2 and SV3 were initially diluted.

Sample SV3 was diluted due to high concentration.

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:51 am, Jul 28, 2025

Signature _____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2617

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/28/2025

Hit Summary Sheet
SW-846

SDG No.: Q2617
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SV1							
Q2617-01	SV1	Air	1,1,1-Trichloroethane	16.4		0.87	1.64	ug/m3
Q2617-01	SV1	Air	Toluene	15.4	J	6.03	18.8	ug/m3
Q2617-01	SV1	Air	Tetrachloroethene	3.87		1.02	2.03	ug/m3
Q2617-01	SV1	Air	Hexane	22.2		5.64	17.6	ug/m3
			Total Voc :	57.9				
			Total Concentration:	57.9				
Client ID:	SV2							
Q2617-02	SV2	Air	Trichloroethene	1.72		1.29	1.61	ug/m3
Q2617-02	SV2	Air	Toluene	14.3	J	6.03	18.8	ug/m3
Q2617-02	SV2	Air	Tetrachloroethene	3.32		1.02	2.03	ug/m3
Q2617-02	SV2	Air	m/p-Xylene	19.1	J	17.8	43.4	ug/m3
Q2617-02	SV2	Air	Hexane	18.3		5.64	17.6	ug/m3
			Total Voc :	56.8				
			Total Concentration:	56.8				
Client ID:	SV3							
Q2617-03	SV3	Air	Heptane	22.1		6.97	20.5	ug/m3
Q2617-03	SV3	Air	1,1,1-Trichloroethane	18.6		0.87	1.64	ug/m3
Q2617-03	SV3	Air	Toluene	39.2		6.03	18.8	ug/m3
Q2617-03	SV3	Air	Tetrachloroethene	16.9		1.02	2.03	ug/m3
Q2617-03	SV3	Air	Ethyl Benzene	384		8.25	21.7	ug/m3
Q2617-03	SV3	Air	m/p-Xylene	1390	E	17.8	43.4	ug/m3
Q2617-03	SV3	Air	o-Xylene	427		9.12	21.7	ug/m3
Q2617-03	SV3	Air	1,3,5-Trimethylbenzene	50.6		8.85	24.6	ug/m3
Q2617-03	SV3	Air	1,2,4-Trimethylbenzene	42.8		8.85	24.6	ug/m3
Q2617-03	SV3	Air	Hexane	95.2		5.64	17.6	ug/m3
			Total Voc :	2490				
			Total Concentration:	2490				
Client ID:	SV3DL							
Q2617-03DL	SV3DL	Air	Heptane	28.7	JD	27.9	82.0	ug/m3
Q2617-03DL	SV3DL	Air	1,1,1-Trichloroethane	19.1	D	3.49	6.55	ug/m3
Q2617-03DL	SV3DL	Air	Toluene	40.0	JD	24.1	75.4	ug/m3
Q2617-03DL	SV3DL	Air	Tetrachloroethene	19.0	D	4.07	8.14	ug/m3
Q2617-03DL	SV3DL	Air	Ethyl Benzene	395	D	33.0	86.9	ug/m3
Q2617-03DL	SV3DL	Air	m/p-Xylene	1610	D	71.2	174	ug/m3
Q2617-03DL	SV3DL	Air	o-Xylene	478	D	36.5	86.9	ug/m3
Q2617-03DL	SV3DL	Air	1,3,5-Trimethylbenzene	48.2	JD	35.4	98.3	ug/m3
Q2617-03DL	SV3DL	Air	1,2,4-Trimethylbenzene	42.3	JD	35.4	98.3	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q2617

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2617-03DL	SV3DL	Air	Hexane	97.3	D	22.6	70.5	ug/m3
			Total Voc :	2770				
			Total Concentration:	2770				

A

B

C

D



SAMPLE DATA

Report of Analysis

Client:	GFE LLC	Date Collected:	07/15/25
Project:	151 Albany Ave, Freeport NY	Date Received:	07/16/25
Client Sample ID:	SV1	SDG No.:	Q2617
Lab Sample ID:	Q2617-01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group2
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042734.D	10		07/22/25 17:07	VL072225

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.25	0.64	U	0.64	0.77	ug/m3
142-82-5	Heptane	1.70	6.97	U	6.97	20.5	ug/m3
75-35-4	1,1-Dichloroethene	1.50	5.95	U	5.95	19.8	ug/m3
110-82-7	Cyclohexane	2.20	7.57	U	7.57	17.2	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	U	3.93	19.8	ug/m3
71-55-6	1,1,1-Trichloroethane	3.00	16.4		0.87	1.64	ug/m3
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	U	6.54	23.4	ug/m3
71-43-2	Benzene	0.79	2.52	U	2.52	16.0	ug/m3
79-01-6	Trichloroethene	0.24	1.29	U	1.29	1.61	ug/m3
108-88-3	Toluene	4.10	15.4	J	6.03	18.8	ug/m3
127-18-4	Tetrachloroethene	0.57	3.87		1.02	2.03	ug/m3
100-41-4	Ethyl Benzene	1.90	8.25	U	8.25	21.7	ug/m3
179601-23-1	m/p-Xylene	4.10	17.8	U	17.8	43.4	ug/m3
95-47-6	o-Xylene	2.10	9.12	U	9.12	21.7	ug/m3
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	U	8.85	24.6	ug/m3
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	U	8.85	24.6	ug/m3
91-20-3	Naphthalene	0.13	0.68	UQ	0.68	5.24	ug/m3
110-54-3	Hexane	6.30	22.2		5.64	17.6	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.4			65 - 135	104%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	110000		2.793			
540-36-3	1,4-Difluorobenzene	321000		3.965			
3114-55-4	Chlorobenzene-d5	281000		8.885			

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	07/15/25
Project:	151 Albany Ave, Freeport NY	Date Received:	07/16/25
Client Sample ID:	SV2	SDG No.:	Q2617
Lab Sample ID:	Q2617-02	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group2
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042736.D	10		07/22/25 18:17	VL072225

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.25	0.64	U	0.64	0.77	ug/m3
142-82-5	Heptane	1.70	6.97	U	6.97	20.5	ug/m3
75-35-4	1,1-Dichloroethene	1.50	5.95	U	5.95	19.8	ug/m3
110-82-7	Cyclohexane	2.20	7.57	U	7.57	17.2	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	U	3.93	19.8	ug/m3
71-55-6	1,1,1-Trichloroethane	0.16	0.87	U	0.87	1.64	ug/m3
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	U	6.54	23.4	ug/m3
71-43-2	Benzene	0.79	2.52	U	2.52	16.0	ug/m3
79-01-6	Trichloroethene	0.32	1.72		1.29	1.61	ug/m3
108-88-3	Toluene	3.80	14.3	J	6.03	18.8	ug/m3
127-18-4	Tetrachloroethene	0.49	3.32		1.02	2.03	ug/m3
100-41-4	Ethyl Benzene	1.90	8.25	U	8.25	21.7	ug/m3
179601-23-1	m/p-Xylene	4.40	19.1	J	17.8	43.4	ug/m3
95-47-6	o-Xylene	2.10	9.12	U	9.12	21.7	ug/m3
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	U	8.85	24.6	ug/m3
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	U	8.85	24.6	ug/m3
91-20-3	Naphthalene	0.13	0.68	UQ	0.68	5.24	ug/m3
110-54-3	Hexane	5.20	18.3		5.64	17.6	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			65 - 135	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	114000		2.807			
540-36-3	1,4-Difluorobenzene	333000		3.978			
3114-55-4	Chlorobenzene-d5	294000		8.888			

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	07/15/25
Project:	151 Albany Ave, Freeport NY	Date Received:	07/16/25
Client Sample ID:	SV3	SDG No.:	Q2617
Lab Sample ID:	Q2617-03	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group2
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042738.D	10		07/22/25 19:26	VL072225

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.25	0.64	U	0.64	0.77	ug/m3
142-82-5	Heptane	5.40	22.1		6.97	20.5	ug/m3
75-35-4	1,1-Dichloroethene	1.50	5.95	U	5.95	19.8	ug/m3
110-82-7	Cyclohexane	2.20	7.57	U	7.57	17.2	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	U	3.93	19.8	ug/m3
71-55-6	1,1,1-Trichloroethane	3.40	18.6		0.87	1.64	ug/m3
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	U	6.54	23.4	ug/m3
71-43-2	Benzene	0.79	2.52	U	2.52	16.0	ug/m3
79-01-6	Trichloroethene	0.24	1.29	U	1.29	1.61	ug/m3
108-88-3	Toluene	10.4	39.2		6.03	18.8	ug/m3
127-18-4	Tetrachloroethene	2.50	16.9		1.02	2.03	ug/m3
100-41-4	Ethyl Benzene	88.4	384		8.25	21.7	ug/m3
179601-23-1	m/p-Xylene	320	1390	E	17.8	43.4	ug/m3
95-47-6	o-Xylene	98.2	427		9.12	21.7	ug/m3
108-67-8	1,3,5-Trimethylbenzene	10.3	50.6		8.85	24.6	ug/m3
95-63-6	1,2,4-Trimethylbenzene	8.70	42.8		8.85	24.6	ug/m3
91-20-3	Naphthalene	0.13	0.68	UQ	0.68	5.24	ug/m3
110-54-3	Hexane	27.0	95.2		5.64	17.6	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.1			65 - 135	101%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	115000			2.803		
540-36-3	1,4-Difluorobenzene	334000			3.972		
3114-55-4	Chlorobenzene-d5	299000			8.885		

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	07/15/25
Project:	151 Albany Ave, Freeport NY	Date Received:	07/16/25
Client Sample ID:	SV3DL	SDG No.:	Q2617
Lab Sample ID:	Q2617-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group2
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042739.D	40		07/22/25 20:01	VL072225

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	1.00	2.56	UD	2.56	3.07	ug/m3
142-82-5	Heptane	7.00	28.7	JD	27.9	82.0	ug/m3
75-35-4	1,1-Dichloroethene	6.00	23.8	UD	23.8	79.3	ug/m3
110-82-7	Cyclohexane	8.80	30.3	UD	30.3	68.8	ug/m3
156-59-2	cis-1,2-Dichloroethene	4.00	15.9	UD	15.9	79.3	ug/m3
71-55-6	1,1,1-Trichloroethane	3.50	19.1	D	3.49	6.55	ug/m3
540-84-1	2,2,4-Trimethylpentane	5.60	26.2	UD	26.2	93.4	ug/m3
71-43-2	Benzene	3.20	10.2	UD	10.2	63.9	ug/m3
79-01-6	Trichloroethene	0.96	5.16	UD	5.16	6.45	ug/m3
108-88-3	Toluene	10.6	40.0	JD	24.1	75.4	ug/m3
127-18-4	Tetrachloroethene	2.80	19.0	D	4.07	8.14	ug/m3
100-41-4	Ethyl Benzene	91.0	395	D	33.0	86.9	ug/m3
179601-23-1	m/p-Xylene	370	1610	D	71.2	174	ug/m3
95-47-6	o-Xylene	110	478	D	36.5	86.9	ug/m3
108-67-8	1,3,5-Trimethylbenzene	9.80	48.2	JD	35.4	98.3	ug/m3
95-63-6	1,2,4-Trimethylbenzene	8.60	42.3	JD	35.4	98.3	ug/m3
91-20-3	Naphthalene	0.52	2.73	UDQ	2.73	21.0	ug/m3
110-54-3	Hexane	27.6	97.3	D	22.6	70.5	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.1			65 - 135	101%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	116000		2.8			
540-36-3	1,4-Difluorobenzene	337000		3.969			
3114-55-4	Chlorobenzene-d5	290000		8.882			

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

LAB CHRONICLE

OrderID:	Q2617	OrderDate:	7/16/2025 11:53:00 AM
Client:	GFE LLC	Project:	151 Albany Ave, Freeport NY
Contact:	Frank Galdun	Location:	--Select--,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2617-01	SV1	Air	VOCMS Group2	TO-15	07/15/25		07/22/25	07/16/25
Q2617-02	SV2	Air	VOCMS Group2	TO-15	07/15/25		07/22/25	07/16/25
Q2617-03	SV3	Air	VOCMS Group2	TO-15	07/15/25		07/22/25	07/16/25
Q2617-03DL	SV3DL	Air	VOCMS Group2	TO-15	07/15/25		07/22/25	07/16/25



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2507015				Courier : FGALDUN				1 of 3 COCs					
Client ID : GFEL01				Project ID : 10-31102-01				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix			
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Individual Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 151 AUBANY AVE FREEPORT, NY													
Analysis Turnaround Time 5 DAY				Standard : 15 Business Days OR				Data Package Type : RESULTS ONLY									
Rush (Specify): 5 Days				EDD Type : PDF													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum In Field ("Hg) (Start)	Can Vacuum In Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SVI	7/14/25	9:45	11:45	30	5.5	72	75	-30	-5.1	10649	10331	6 L	50	VL042564.D	1		1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15)							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY THOSE ANALYTES ON THE ATTACHED LIST. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 0.0																	
Sampling site (State):																	
Quick Connector required : No																	
Canisters Shipped by: FG				Date/Time: 7/14/25				Canisters Received by: CR				Date/Time: 7/16/25 11:05					
Samples Relinquished by: FG				Date/Time: 7/16/25				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					
B2507015 - 1																	

Client Contact Information				Bottle Order ID : B2507014				Courier : F GALDUN				2 of 3 COCs				
Client ID : GFEL01				Project ID : 151 ALBANY AVE, NY				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix		
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified								
				Phone Number : 646-542-3465												
				Fax Number : 973-334-1692												
				Site Details: 151 ALBANY AVE FREEDPORT, NY												
				Analysis Turnaround Time 5 DAY				Data Package Type RESULTS ONLY								
				Standard : 10 business days OR				EDD Type : PDF								
				Rush (Specify): 5 Days												
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SN2	7/16/25	10:40	12:40	30	4	72	75	-30	-3.9	10226	10283	6 L	50	VL042600.D	1	1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 						
		Ambient	Maximum	Minimum												
Start																
Stop																
Pressure (Inches of Hg)										** Submittal of this COC Indicates approval of the analysis based on existing conditions. REPORT ONLY THOSE ANALYTES ON THE ATTACHED LIST Please follow the instructions on the back of this COC.						
		Ambient	Maximum	Minimum												
Start																
Stop																
Special Instructions/QC Requirements & Comments :																
Suspected Contamination: High Medium <u>Low</u> PID Readings: 0.0																
Sampling site (State):																
Quick Connector required : NO																
Canisters Shipped by:				Date/Time: 07/14/25				Canisters Received by:				Date/Time: 7/16/25 11:05				
Samples Relinquished by:				Date/Time: 7/16/25				Received by:				Date/Time:				
Relinquished by:				Date/Time:				Received by:				Date/Time:				
																B2507014 - 1

Client Contact Information				Bottle Order ID : B2507015				Courier : FGA22WJ				3 of 3 COCs																
Client ID : GFEL01				Project ID : 10 Elridge St.				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix														
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																				
				Phone Number : 646-542-3465																								
				Fax Number : 973-334-1692																								
				Site Details: 151 ALBANY AVE FREETPORT, NY																								
Analysis Turnaround Time 5 DAY				Standard : 10 business days OR				Data Package Type : RESULTS ONLY				Indoor/Ambient Air		Soil Gas														
Rush (Specify): 5 Days				EDD Type : EDF																								
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15													
SV3	7/16/25	11:55	1:55	30	6	12	75	-30	-6.7	10579	10608	6 L	50	VL042564.D	1	1												
Temperature (Fahrenheit) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) Sut						
	Ambient	Maximum	Minimum																									
Start																												
Stop																												
Pressure (Inches of Hg) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY THOSE ANALYTES ON THE ATTACHED LIST. Please follow the instructions on the back of this COC.						
	Ambient	Maximum	Minimum																									
Start																												
Stop																												
Special Instructions/QC Requirements & Comments :																												
Suspected Contamination: High Medium Low PID Readings: 0.0																												
Sampling site (State):																												
Quick Connector required : NO																												
Canisters Shipped by: SAM				Date/Time: 7/16/25 14				Canisters Received by: CL				Date/Time: 7/16/25 11:5																
Samples Relinquished by: FGA				Date/Time: 7/16/25				Received by:				Date/Time:																
Relinquished by:				Date/Time:				Received by:				Date/Time:																
																B2507015 - 2												

REQUESTED ANALYTE LIST:

PCE *Tetrachloroethene*

TCE *Trichloroethene*

cis-1,2-DCE *ene*

1,1,1-TCA *ane*

1,1-DCE *ene*

Vinyl chloride

Benzene

Toluene

Ethylbenzene

Naphthalene

Cyclohexane

2,2,4-Trimethylpentane

1,2,4-Trimethylbenzene

1,3,5-Trimethylbenzene

o-xylene

m,p-xylene

Heptane

Hexane

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample

Laboratory: Chemtech

Location: 284 Sheffield Street, Mountainside, NJ 7092

NA66:

Title: Sample Custodian

Field Sample Seal No.: Q2617



Date Broken: 7/16/2025

Military Time Seal Broken: 11:05:00

Case No.: 151 Albany Ave, Freeport I

Analytical Parameter/Fraction/OCMS Group: 2

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q2617-01	SV1		
Q2617-02	SV2		
Q2617-03	SV3		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
7/16/25	12:15	Signature 	Signature 	
		Printed Name <u>Cassanova Pera</u>	Printed Name <u>Sam S. H. Yen</u>	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy