

DATA PACKAGE

GC SEMI-VOLATILES
VOLATILE ORGANICS

PROJECT NAME : 166 BELMONT AVE BELLEVILLE

SCIACCA GENERAL CONTRACTORS, LLC

2 Shaw Court

Fairfield, NJ - 07004

Phone No: 201-933-6100

ORDER ID : Q2618

ATTENTION : Rosanne Scirica



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOC-TCLVOA-10- Case Narrative	4
2.2) TPH GC- Case Narrative	6
2.3) EPH_F2- Case Narrative	8
3) Qualifier Page	10
4) QA Checklist	11
5) VOC-TCLVOA-10 Data	12
6) TPH GC Data	18
7) EPH_F2 Data	21
8) Shipping Document	41
8.1) CHAIN OF CUSTODY	42
8.2) Lab Certificate	43
8.3) Internal COC	44

1
2
3
4
5
6
7
8

Cover Page

Order ID : Q2618

Project ID : 166 Belmont Ave Belleville

Client : Sciacca General Contractors, LLC

Lab Sample Number

Q2618-01
Q2618-02
Q2618-03
Q2618-04
Q2618-05
Q2618-06
Q2618-07
Q2618-08

Client Sample Number

WASTE
VOC
1
2
3
4
5
6

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/26/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Sciacca General Contractors, LLC Project

Name: 166 Belmont Ave Belleville

Project # N/A

Order ID # Q2618

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 07/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_W were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration method (82W063025S.M) for Methylene Chloride passing on Quadratic Regression.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Sciacca General Contractors, LLC

Project Name: 166 Belmont Ave Belleville

Project # N/A

Order ID # Q2618

Test Name: TPH GC

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 07/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:
TPH GC. This data package contains results for TPH GC.

C. Analytical Techniques:

The analysis were performed on instrument FID_G. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 13302. The analysis of TPH GC was based on method 8015D and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q2592-01MS} with File ID: FG016339.D recoveries met the requirements for all compounds except for Petroleum Hydrocarbons[2163%] due to matrix interference.

The MSD {Q2592-01MSD} with File ID: FG016340.D recoveries met the requirements for all compounds except for Petroleum Hydrocarbons[2106%] due to matrix interference.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

Samples WASTE was diluted due to bad matrix.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Sciacca General Contractors, LLC

Project Name: 166 Belmont Ave Belleville

Project # N/A

Order ID # Q2618

Test Name: EPH_F2

A. Number of Samples and Date of Receipt:

6 Solid samples were received on 07/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH_F2. This data package contains results for EPH_F2.

C. Analytical Techniques:

The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of EPH_F2s was based on method NJEPH and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q2618-08MS} with File ID: FE054930.D recoveries met the requirements for all compounds except for Aliphatic [n-Decane (C10)- 146%] due to matrix interference.

The MS {Q2618-08MSD} with File ID: FE054931.D recoveries met the requirements for all compounds except for Aliphatic [n-Decane (C10)- 141%] due to matrix interference.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: <ol style="list-style-type: none"> (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2618

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/26/2025

Hit Summary Sheet
SW-846

SDG No.: Q2618

Client: Sciacca General Contractors, LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



SAMPLE DATA

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/16/25	
Project:	166 Belmont Ave Belleville		Date Received:	07/16/25	
Client Sample ID:	VOC		SDG No.:	Q2618	
Lab Sample ID:	Q2618-02		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	92.9	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW031874.D	1	07/16/25 14:53	VW071625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.20	U	1.20	5.40	ug/Kg
74-87-3	Chloromethane	1.20	U	1.20	5.40	ug/Kg
75-01-4	Vinyl Chloride	0.85	U	0.85	5.40	ug/Kg
74-83-9	Bromomethane	1.20	U	1.20	5.40	ug/Kg
75-00-3	Chloroethane	1.40	U	1.40	5.40	ug/Kg
75-69-4	Trichlorofluoromethane	1.30	U	1.30	5.40	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.40	ug/Kg
75-35-4	1,1-Dichloroethene	1.10	U	1.10	5.40	ug/Kg
67-64-1	Acetone	5.10	U	5.10	26.9	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.40	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.79	U	0.79	5.40	ug/Kg
79-20-9	Methyl Acetate	1.70	U	1.70	5.40	ug/Kg
75-09-2	Methylene Chloride	3.80	U	3.80	10.8	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.93	U	0.93	5.40	ug/Kg
75-34-3	1,1-Dichloroethane	0.86	U	0.86	5.40	ug/Kg
110-82-7	Cyclohexane	0.85	U	0.85	5.40	ug/Kg
78-93-3	2-Butanone	7.00	U	7.00	26.9	ug/Kg
56-23-5	Carbon Tetrachloride	1.00	U	1.00	5.40	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.81	U	0.81	5.40	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.40	ug/Kg
67-66-3	Chloroform	0.90	U	0.90	5.40	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.00	U	1.00	5.40	ug/Kg
108-87-2	Methylcyclohexane	0.98	U	0.98	5.40	ug/Kg
71-43-2	Benzene	0.85	U	0.85	5.40	ug/Kg
107-06-2	1,2-Dichloroethane	0.85	U	0.85	5.40	ug/Kg
79-01-6	Trichloroethene	0.87	U	0.87	5.40	ug/Kg
78-87-5	1,2-Dichloropropane	0.98	U	0.98	5.40	ug/Kg
75-27-4	Bromodichloromethane	0.84	U	0.84	5.40	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.90	U	3.90	26.9	ug/Kg
108-88-3	Toluene	0.84	U	0.84	5.40	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/16/25	
Project:	166 Belmont Ave Belleville		Date Received:	07/16/25	
Client Sample ID:	VOC		SDG No.:	Q2618	
Lab Sample ID:	Q2618-02		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	92.9	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW031874.D	1	07/16/25 14:53	VW071625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.70	U	0.70	5.40	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.67	U	0.67	5.40	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.99	U	0.99	5.40	ug/Kg
591-78-6	2-Hexanone	4.00	U	4.00	26.9	ug/Kg
124-48-1	Dibromochloromethane	0.94	U	0.94	5.40	ug/Kg
106-93-4	1,2-Dibromoethane	0.95	U	0.95	5.40	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.40	ug/Kg
108-90-7	Chlorobenzene	0.98	U	0.98	5.40	ug/Kg
100-41-4	Ethyl Benzene	0.72	U	0.72	5.40	ug/Kg
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.8	ug/Kg
95-47-6	o-Xylene	0.88	U	0.88	5.40	ug/Kg
100-42-5	Styrene	0.76	U	0.76	5.40	ug/Kg
75-25-2	Bromoform	0.93	U	0.93	5.40	ug/Kg
98-82-8	Isopropylbenzene	0.84	U	0.84	5.40	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.30	U	1.30	5.40	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.80	U	1.80	5.40	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.70	U	1.70	5.40	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.60	U	1.60	5.40	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	2.00	5.40	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.20	U	3.20	5.40	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.40	U	3.40	5.40	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.6		63 - 155	95%	SPK: 50
1868-53-7	Dibromofluoromethane	45.7		70 - 134	91%	SPK: 50
2037-26-5	Toluene-d8	44.2		74 - 123	88%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.7		17 - 146	93%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	170000	7.965			
540-36-3	1,4-Difluorobenzene	346000	8.849			
3114-55-4	Chlorobenzene-d5	314000	11.629			
3855-82-1	1,4-Dichlorobenzene-d4	140000	13.556			

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/16/25	
Project:	166 Belmont Ave Belleville		Date Received:	07/16/25	
Client Sample ID:	VOC		SDG No.:	Q2618	
Lab Sample ID:	Q2618-02		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	92.9	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW031874.D	1	07/16/25 14:53	VW071625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q2618	OrderDate:	7/16/2025 12:13:00 PM
Client:	Sciacca General Contractors, LLC	Project:	166 Belmont Ave Belleville
Contact:	Rosanne Scirica	Location:	--Select--,O11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2618-02	VOC	SOIL	VOC-TCLVOA-10	8260D	07/16/25		07/16/25	07/16/25



SAMPLE DATA

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	WASTE	SDG No.:	Q2618
Lab Sample ID:	Q2618-01	Matrix:	SOIL
Analytical Method:	8015D TPH	% Solid:	94.5
Sample Wt/Vol:	30.1	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TPH GC
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG016334.D	5	07/17/25 08:25	07/18/25 11:54	PB168895

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
PHC	Petroleum Hydrocarbons	66900		2030	14900	ug/kg
SURROGATES						
16416-32-3	TETRACOSANE-d50	2.34		37 - 130	58%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2618	OrderDate:	7/16/2025 12:13:00 PM
Client:	Sciacca General Contractors, LLC	Project:	166 Belmont Ave Belleville
Contact:	Rosanne Scirica	Location:	--Select--,O11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2618-01	WASTE	SOIL			07/16/25			07/16/25
			TPH GC	8015D		07/17/25	07/18/25	
Q2618-03	1	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	
Q2618-04	2	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	
Q2618-05	3	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	
Q2618-06	4	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	
Q2618-07	5	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	
Q2618-08	6	Solid			07/16/25			07/16/25
			EPH_F2	NJEPH		07/18/25	07/18/25	



SAMPLE DATA

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	1	SDG No.:	Q2618
Lab Sample ID:	Q2618-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	91.1
Sample Wt/Vol:	30.03 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 13:55	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	11.8		1	1.00	4.39	mg/kg	FE054924.D
Total EPH	Total EPH	11.8			1.00	4.39	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	1	SDG No.:	Q2618
Lab Sample ID:	Q2618-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	91.1
Sample Wt/Vol:	30.03 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054924.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	11.8		1.00	4.39	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	4.77		1.29	2.19	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	33.7		40 - 140	67%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	34.0		40 - 140	68%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-03	Acq On:	18 Jul 2025 13:55
Client Sample ID:	1	Operator:	YP\AJ
Data file:	FE054924.D	Misc:	
Instrument:	FID_E	ALS Vial:	14
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	296089	2.176	300	ug/ml
Aliphatic C12-C16	6.964	10.415	1322504	9.408	200	ug/ml
Aliphatic C16-C21	10.416	13.793	15154465	104.952	300	ug/ml
Aliphatic C21-C28	13.794	17.463	6548171	45.269	400	ug/ml
Aliphatic C28-C40	17.464	22.490	9040287	65.203	600	ug/ml
Aliphatic EPH	3.330	22.490	32361516	227.009		ug/ml
ortho-Terphenyl (SURR)	12.088	12.088	5520080	33.99		ug/ml
1-chlorooctadecane (SURR)	13.524	13.524	4260430	33.73		ug/ml
Aliphatic C9-C28	3.330	17.463	23321229	161.805	1200	ug/ml

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	2	SDG No.:	Q2618
Lab Sample ID:	Q2618-04	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.6
Sample Wt/Vol:	30.08 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 14:25	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	38.7		1	0.97	4.27	mg/kg	FE054925.D
Total EPH	Total EPH	38.7			0.97	4.27	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	2	SDG No.:	Q2618
Lab Sample ID:	Q2618-04	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.6
Sample Wt/Vol:	30.08 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054925.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	38.7		0.97	4.27	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	4.81		1.26	2.13	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	33.2		40 - 140	66%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	34.0		40 - 140	68%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-04	Acq On:	18 Jul 2025 14:25
Client Sample ID:	2	Operator:	YP\AJ
Data file:	FE054925.D	Misc:	
Instrument:	FID_E	ALS Vial:	15
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	1836878	13.501	300	ug/ml
Aliphatic C12-C16	6.964	10.415	5348353	38.048	200	ug/ml
Aliphatic C16-C21	10.416	13.793	51986094	360.03	300	ug/ml
Aliphatic C21-C28	13.794	17.463	19216433	132.846	400	ug/ml
Aliphatic C28-C40	17.464	22.490	9397576	67.78	600	ug/ml
Aliphatic EPH	3.330	22.490	87785334	612.206		ug/ml
ortho-Terphenyl (SURR)	12.088	12.088	5519565	33.99		ug/ml
1-chlorooctadecane (SURR)	13.524	13.524	4191336	33.19		ug/ml
Aliphatic C9-C28	3.330	17.463	78387758	544.425	1200	ug/ml

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	3	SDG No.:	Q2618
Lab Sample ID:	Q2618-05	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	94.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 14:55	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	40.5		1	0.97	4.24	mg/kg	FE054926.D
Total EPH	Total EPH	40.5			0.97	4.24	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	3	SDG No.:	Q2618
Lab Sample ID:	Q2618-05	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	94.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054926.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	40.5		0.97	4.24	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	4.46		1.25	2.12	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	38.2		40 - 140	76%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	39.2		40 - 140	78%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-05	Acq On:	18 Jul 2025 14:55
Client Sample ID:	3	Operator:	YP\AJ
Data file:	FE054926.D	Misc:	
Instrument:	FID_E	ALS Vial:	16
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	538939	3.961	300	ug/ml
Aliphatic C12-C16	6.964	10.415	5705176	40.587	200	ug/ml
Aliphatic C16-C21	10.416	13.793	55671465	385.553	300	ug/ml
Aliphatic C21-C28	13.794	17.463	20903504	144.509	400	ug/ml
Aliphatic C28-C40	17.464	22.490	8743883	63.066	600	ug/ml
Aliphatic EPH	3.330	22.490	91562967	637.676		ug/ml
ortho-Terphenyl (SURR)	12.089	12.089	6359911	39.16		ug/ml
1-chlorooctadecane (SURR)	13.524	13.524	4821724	38.18		ug/ml
Aliphatic C9-C28	3.330	17.463	82819084	574.61	1200	ug/ml

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	4	SDG No.:	Q2618
Lab Sample ID:	Q2618-06	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.4
Sample Wt/Vol:	30.04 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 15:25	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	36.1		1	0.97	4.28	mg/kg	FE054927.D
Total EPH	Total EPH	36.1			0.97	4.28	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	4	SDG No.:	Q2618
Lab Sample ID:	Q2618-06	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.4
Sample Wt/Vol:	30.04 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054927.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	36.1		0.97	4.28	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	4.78		1.26	2.14	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	35.2		40 - 140	70%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	36.2		40 - 140	72%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-06	Acq On:	18 Jul 2025 15:25
Client Sample ID:	4	Operator:	YP\AJ
Data file:	FE054927.D	Misc:	
Instrument:	FID_E	ALS Vial:	17
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	626015	4.601	300	ug/ml
Aliphatic C12-C16	6.964	10.415	4525977	32.198	200	ug/ml
Aliphatic C16-C21	10.416	13.793	48703538	337.297	300	ug/ml
Aliphatic C21-C28	13.794	17.463	19184881	132.628	400	ug/ml
Aliphatic C28-C40	17.464	22.490	9300394	67.079	600	ug/ml
Aliphatic EPH	3.330	22.490	82340805	573.803		ug/ml
ortho-Terphenyl (SURR)	12.088	12.088	5874568	36.17		ug/ml
1-chlorooctadecane (SURR)	13.523	13.523	4441371	35.17		ug/ml
Aliphatic C9-C28	3.330	17.463	73040411	506.724	1200	ug/ml

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	5	SDG No.:	Q2618
Lab Sample ID:	Q2618-07	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.8
Sample Wt/Vol:	30.02 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 15:56	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	27.5		1	0.97	4.27	mg/kg	FE054928.D
Total EPH	Total EPH	27.5			0.97	4.27	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	5	SDG No.:	Q2618
Lab Sample ID:	Q2618-07	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.8
Sample Wt/Vol:	30.02 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054928.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	27.5		0.97	4.27	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	5.03		1.26	2.13	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	36.0		40 - 140	72%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	36.7		40 - 140	73%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-07	Acq On:	18 Jul 2025 15:56
Client Sample ID:	5	Operator:	YP\AJ
Data file:	FE054928.D	Misc:	
Instrument:	FID_E	ALS Vial:	18
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	413575	3.04	300	ug/ml
Aliphatic C12-C16	6.964	10.415	3227155	22.958	200	ug/ml
Aliphatic C16-C21	10.416	13.793	37667311	260.865	300	ug/ml
Aliphatic C21-C28	13.794	17.463	14606713	100.978	400	ug/ml
Aliphatic C28-C40	17.464	22.490	9823909	70.855	600	ug/ml
Aliphatic EPH	3.330	22.490	65738663	458.697		ug/ml
ortho-Terphenyl (SURR)	12.088	12.088	5955040	36.67		ug/ml
1-chlorooctadecane (SURR)	13.524	13.524	4552503	36.05		ug/ml
Aliphatic C9-C28	3.330	17.463	55914754	387.841	1200	ug/ml

Report of Analysis

Client:	Sciaccia General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	6	SDG No.:	Q2618
Lab Sample ID:	Q2618-08	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.4
Sample Wt/Vol:	30.09 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
07/18/25 08:55	07/18/25 16:26	PB168910

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C9-C28	Aliphatic C9-C28	20.3		1	0.97	4.27	mg/kg	FE054929.D
Total EPH	Total EPH	20.3			0.97	4.27	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/16/25
Project:	166 Belmont Ave Belleville	Date Received:	07/16/25
Client Sample ID:	6	SDG No.:	Q2618
Lab Sample ID:	Q2618-08	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	93.4
Sample Wt/Vol:	30.09 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_F2
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE054929.D	1	07/18/25	07/18/25	PB168910

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	20.3		0.97	4.27	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	4.84		1.26	2.13	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	35.5		40 - 140	71%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	35.5		40 - 140	71%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q2618-08	Acq On:	18 Jul 2025 16:26
Client Sample ID:	6	Operator:	YP\AJ
Data file:	FE054929.D	Misc:	
Instrument:	FID_E	ALS Vial:	19
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.330	6.963	250876	1.844	300	ug/ml
Aliphatic C12-C16	6.964	10.415	5398377	38.404	200	ug/ml
Aliphatic C16-C21	10.416	13.793	27157696	188.081	300	ug/ml
Aliphatic C21-C28	13.794	17.463	8373718	57.889	400	ug/ml
Aliphatic C28-C40	17.464	22.490	9426343	67.988	600	ug/ml
Aliphatic EPH	3.330	22.490	50607010	354.206		ug/ml
ortho-Terphenyl (SURR)	12.088	12.088	5770269	35.53		ug/ml
1-chlorooctadecane (SURR)	13.524	13.524	4483088	35.5		ug/ml
Aliphatic C9-C28	3.330	17.463	41180667	286.218	1200	ug/ml

LAB CHRONICLE

OrderID:	Q2618	OrderDate:	7/16/2025 12:13:00 PM
Client:	Sciacca General Contractors, LLC	Project:	166 Belmont Ave Belleville
Contact:	Rosanne Scirica	Location:	--Select--,O11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2618-01	WASTE	SOIL	TPH GC	8015D	07/16/25	07/17/25	07/18/25	07/16/25
Q2618-03	1	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25
Q2618-04	2	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25
Q2618-05	3	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25
Q2618-06	4	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25
Q2618-07	5	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25
Q2618-08	6	Solid	EPH_F2	NJEPH	07/16/25	07/18/25	07/18/25	07/16/25



SHIPPING DOCUMENTS

166 Belmont Ave
 Belleville.
 166 Belmont Ave Q2618

CHEMTECH
 CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
 (908) 789-8900 Fax (908) 789-8922
 www.chemtech.net

Chemtech Project Number Belleville
 COC Number

CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION	
Report to be sent to:		PROJECT NAME:		BILL TO:	
COMPANY:		PROJECT #:		PO#	
ADDRESS:		LOCATION:		ADDRESS:	
CITY:		PROJECT MANAGER:		CITY:	
STATE:		E-MAIL:		STATE:	
ATTENTION:		PHONE:		ATTENTION:	
PHONE:		FAX:		PHONE:	
ZIP:				ZIP:	

DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS	
FAX (RUSH) _____ DAYS*		<input type="checkbox"/> Level 1 (Results Only) <input type="checkbox"/> Level 4 (QC + Full Raw Data) <input type="checkbox"/> Level 2 (Results + QC) <input type="checkbox"/> NJ Reduced <input type="checkbox"/> US EPA CLP <input type="checkbox"/> Level 3 (Results + QC + Raw Data) <input type="checkbox"/> NYS ASP A <input type="checkbox"/> NYS ASP B <input type="checkbox"/> EDD FORMAT <input type="checkbox"/> Other _____		<div style="border: 1px solid black; padding: 5px;"> TPH-GC VOC EPH-F2 </div>	
HARDCOPY (DATA PACKAGE): _____ DAYS*					
EDD: _____ DAYS*					
TO BE APPROVED BY CHEMTECH					
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS					

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	WASTE				7/16/25	8:30	1	X										
2.	VOC				7/16/25	8:30	1		X									
3.	1				7/16/25	8:45	1			X								
4.	2				7/16/25	8:45	1			X								
5.	3				7/16/25	9:05	1			X								
6.	4				7/16/25	10:00	1			X								
7.	5				7/16/25	10:30	1			X								
8.	6				7/16/25	9:30	1			X								
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME 1145 7-16-25	RECEIVED BY 1. <u>[Signature]</u> 7-16-25	Conditions of bottles or collars at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP <u>5.15</u>
RELINQUISHED BY	DATE/TIME	RECEIVED BY	Comments:
RELINQUISHED BY	DATE/TIME 1318 7-16-25	RECEIVED FOR LAB BY 3. <u>[Signature]</u>	CLIENT: <input type="checkbox"/> Hand Delivered <input type="checkbox"/> Other: _____ CHEMTECH: <input type="checkbox"/> Picked Up Shipment Complete <input type="checkbox"/> YES <input type="checkbox"/> NO

10/2018 WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2618	SCIA01	Order Date : 7/16/2025 12:13:00 PM	Project Mgr : Deepak
Client Name : Sciacca General Contractor		Project Name : 98 Morse Ave Nutley 166 Belmont Ave NJ	Report Type : Results Only
Client Contact : Rosanne Scirica		Receive DateTime : 7/16/2025 2:00:00 PM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Sciacca General Contractor		Purchase Order : 13:151M	Hard Copy Date :
Invoice Contact : Rosanne Scirica			Date Signoff : 7/16/2025 1:55:47 PM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2618-02	VOC	Solid	07/16/2025	08:30	VOC-TCLVOA-10		8260D		10 Bus. Days


Relinquished By :

Date / Time :


7-16-25 1420

Received By :

Date / Time :


07/16/25 14:20

Storage Area : VOA Refridgerator Room

Ref # 6
E 22