

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG No.: Q2618
 Instrument ID: MSVOA_W Calibration Date/Time: 07/16/2025 09:22
 Lab File ID: VW031864.D Init. Calib. Date(s): 06/30/2025 06/30/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:54 12:55
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	54.07	-8.134003	20