

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW063025

Review By	Maresh Dadoda	Review On	7/1/2025 12:21:51 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/1/2025 12:24:32 PM
SubDirectory	VW063025	HP Acquire Method	MSVOA_W
		HP Processing Method	82w063025s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134578
Initial Calibration Stds	VP134581,VP134582,VP134583,VP134584,VP134585,VP134586
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031728.D	30 Jun 2025 08:56		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW031729.D	30 Jun 2025 09:54		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW031730.D	30 Jun 2025 10:15	8260-soil-method	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW031731.D	30 Jun 2025 10:58		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW031732.D	30 Jun 2025 11:21	Comp. #20 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW031733.D	30 Jun 2025 12:34		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW031734.D	30 Jun 2025 12:55		SY/MD	Ok,M
8	VIBLK	VIBLK	VW031735.D	30 Jun 2025 14:11		SY/MD	Ok
9	VSTDICV050	ICVVW063025	VW031736.D	30 Jun 2025 15:23		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW071625

Review By	Maresh Dadoda	Review On	7/17/2025 8:18:11 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:27:26 AM
SubDirectory	VW071625	HP Acquire Method	MSVOA_W HP Processing Method 82w063025s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134771
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134772,VP134773 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031863.D	16 Jul 2025 08:53		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW031864.D	16 Jul 2025 09:22		SY/MD	Ok,M
3	VW0716SBL01	VW0716SBL01	VW031865.D	16 Jul 2025 09:52		SY/MD	Ok
4	VW0716SBS01	VW0716SBS01	VW031866.D	16 Jul 2025 10:22		SY/MD	Ok,M
5	VW0716SBSD01	VW0716SBSD01	VW031867.D	16 Jul 2025 10:44		SY/MD	Ok,M
6	Q2612-01	OR-02-071525	VW031868.D	16 Jul 2025 11:23	Internal standard fail;Surrogate fail,vial-A	SY/MD	ReRun
7	Q2611-01	EO-02-071525	VW031869.D	16 Jul 2025 11:45	vial A	SY/MD	Ok
8	Q2614-02	HR-MCN-VOC-01	VW031870.D	16 Jul 2025 12:07	vial A	SY/MD	Ok
9	Q2609-03	710-C	VW031871.D	16 Jul 2025 12:29	vial A Need MeOH	SY/MD	Dilution
10	Q2609-07	709-A	VW031872.D	16 Jul 2025 12:51	vial A Internal Standard Fail;Need MeOH	SY/MD	Dilution
11	VIBLK	VIBLK	VW031873.D	16 Jul 2025 13:13		SY/MD	Ok
12	Q2618-02	VOC	VW031874.D	16 Jul 2025 14:53	vial A	SY/MD	Ok
13	Q2622-01	2819	VW031875.D	16 Jul 2025 15:15	vial A	SY/MD	Ok
14	VIBLK	VIBLK	VW031876.D	16 Jul 2025 15:38		SY/MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VW031877.D	16 Jul 2025 16:00		SY/MD	Ok,M

M : Manual Integration