

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2618
 Lab File ID: VW031864.D Date Analyzed: 07/16/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	223578	7.96	416969	8.85	361888	11.64
UPPER LIMIT	447156	8.459	833938	9.349	723776	12.135
LOWER LIMIT	111789	7.459	208485	8.349	180944	11.135
EPA SAMPLE NO.						
VOC	170339	7.97	346398	8.85	314446	11.63
VW0716SBL01	185902	7.97	374504	8.85	352304	11.63
VW0716SBS01	220255	7.96	408210	8.85	372511	11.64
VW0716SBSD01	214931	7.97	411558	8.86	360281	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.