

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2618
 Lab File ID: VW031864.D Date Analyzed: 07/16/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	164563	13.556				
UPPER LIMIT	329126	14.056				
LOWER LIMIT	82281.5	13.056				
EPA SAMPLE NO.						
VOC	139848	13.56				
VW0716SBL01	163773	13.56				
VW0716SBS01	173720	13.56				
VW0716SBSD01	174902	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.