



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q2629

**Order ID:** Q2629

**Client:** Tully Environmental, Inc

**Project ID:** Transfer Station-SPDES

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
<b>Client ID :</b> Q2629-01	<b>001 willets Pt Blvd(may)</b> 001 willets Pt Blvd(may)	Water	Iron	7150		18.0	50.0	ug/L
<b>Client ID :</b> Q2629-02	<b>002 35th Ave(may)</b> 002 35th Ave(may)	Water	Iron	10400		18.0	50.0	ug/L



# SAMPLE DATA

A

B

C

D

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	07/16/25
Project:	Transfer Station-SPDES	Date Received:	07/17/25
Client Sample ID:	001 willets Pt Blvd(may)	SDG No.:	Q2629
Lab Sample ID:	Q2629-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	7150	1	18.0		50.0	ug/L	07/17/25 12:05	07/18/25 12:48	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 12			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	07/16/25
Project:	Transfer Station-SPDES	Date Received:	07/17/25
Client Sample ID:	002 35th Ave(may)	SDG No.:	Q2629
Lab Sample ID:	Q2629-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	10400	1	18.0		50.0	ug/L	07/17/25 12:05	07/18/25 12:52	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 12			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## LAB CHRONICLE

<b>OrderID:</b>	Q2629	<b>OrderDate:</b>	7/17/2025 10:41:00 AM
<b>Client:</b>	Tully Environmental, Inc	<b>Project:</b>	Transfer Station-SPDES
<b>Contact:</b>	Dean Devoe	<b>Location:</b>	O32

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2629-01</b>	<b>001 willets Pt Blvd(may)</b>	<b>Water</b>			<b>07/16/25</b>			<b>07/17/25</b>
			Metals Group 12	200.7		07/17/25	07/18/25	
<b>Q2629-02</b>	<b>002 35th Ave(may)</b>	<b>Water</b>			<b>07/16/25</b>			<b>07/17/25</b>
			Metals Group 12	200.7		07/17/25	07/18/25	



A

B

C

# SAMPLE DATA

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	07/16/25 12:15
Project:	Transfer Station-SPDES	Date Received:	07/17/25
Client Sample ID:	001 willets Pt Blvd(may)	SDG No.:	Q2629
Lab Sample ID:	Q2629-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TSS	207		1	1.00	4.00	mg/L		07/17/25 15:00	SM 2540 D-20

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	07/16/25 12:15
Project:	Transfer Station-SPDES	Date Received:	07/17/25
Client Sample ID:	002 35th Ave(may)	SDG No.:	Q2629
Lab Sample ID:	Q2629-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TSS	191		1	1.00	4.00	mg/L		07/17/25 15:00	SM 2540 D-20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



## LAB CHRONICLE

<b>OrderID:</b>	Q2629	<b>OrderDate:</b>	7/17/2025 10:41:00 AM
<b>Client:</b>	Tully Environmental, Inc	<b>Project:</b>	Transfer Station-SPDES
<b>Contact:</b>	Dean Devoe	<b>Location:</b>	O32

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2629-01	001 willets Pt Blvd(may)	WATER	TSS	SM2540 D	07/16/25 12:15		07/17/25 15:00	07/17/25
Q2629-02	002 35th Ave(may)	WATER	TSS	SM2540 D	07/16/25 12:15		07/17/25 15:00	07/17/25