



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Tully Environmental, Inc**

**Project Name: Transfer Station-SPDES**

**Project # N/A**

**Order ID # Q2630**

**Test Name: TSS**

**A. Number of Samples and Date of Receipt:**

2 Water samples were received on 07/17/2025.

**B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: TSS. This data package contains results for TSS.

**C. Analytical Techniques:**

The analysis of TSS was based on method SM2540 D.

**D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

**E. Additional Comments:**

The temperature of the samples at the time of receipt was 20.5°C, Lab notified this issue to the client. See the communication in shipping Document section.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_