

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136546

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/18/25 11:19		Jaswal	OK
2	S1	S1	CAL2	07/18/25 11:24		Jaswal	OK
3	S2	S2	CAL3	07/18/25 11:28		Jaswal	OK
4	S3	S3	CAL4	07/18/25 11:32		Jaswal	OK
5	S4	S4	CAL5	07/18/25 11:36		Jaswal	OK
6	S5	S5	CAL6	07/18/25 11:40		Jaswal	OK
7	ICV01	ICV01	ICV	07/18/25 12:01	ICV fail for Cr,Na,TI (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/18/25 12:05		Jaswal	OK
9	ICB01	ICB01	ICB	07/18/25 12:10		Jaswal	OK
10	CRI01	CRI01	CRDL	07/18/25 12:14		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/18/25 12:18		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/18/25 12:23		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/18/25 12:28		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/18/25 12:32		Jaswal	OK
15	CCV01	CCV01	CCV	07/18/25 12:36		Jaswal	OK
16	CCB01	CCB01	CCB	07/18/25 12:43		Jaswal	OK
17	Q2629-01	011 willets Pt Blvd(ma	SAM	07/18/25 12:48		Jaswal	OK
18	Q2629-02	022 35th Ave(may)	SAM	07/18/25 12:52		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136546

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	Q2630-01	001 Willets Pt Blvd (JUNE)	SAM	07/18/25 12:56		Jaswal	OK
20	Q2630-02	002 35th Ave (JUNE)	SAM	07/18/25 13:00		Jaswal	OK
21	PB168908BL	PB168908BL	MB	07/18/25 13:04		Jaswal	OK
22	PB168908BS	PB168908BS	LCS	07/18/25 13:08		Jaswal	OK
23	Q2630-02DUP	002 35th Ave (JUNE)	DUP	07/18/25 13:12		Jaswal	OK
24	Q2630-02L	002 35th Ave (JUNE)	SD	07/18/25 13:16		Jaswal	OK
25	Q2630-02MS	002 35th Ave (JUNE)	MS	07/18/25 13:21		Jaswal	OK
26	Q2630-02MSD	002 35th Ave (JUNE)	MSD	07/18/25 13:25		Jaswal	OK
27	CCV02	CCV02	CCV	07/18/25 13:29		Jaswal	OK
28	CCB02	CCB02	CCB	07/18/25 13:33		Jaswal	OK
29	Q2630-02A	002 35th Ave (JUNE)	PS	07/18/25 13:37		Jaswal	OK
30	PB168901BL	PB168901BL	MB	07/18/25 13:41		Jaswal	OK
31	Q2623-01	3362	SAM	07/18/25 13:49		Jaswal	OK
32	Q2623-02	3363	SAM	07/18/25 13:53		Jaswal	OK
33	Q2624-01	449	SAM	07/18/25 13:58		Jaswal	OK
34	Q2624-02	450	SAM	07/18/25 14:02		Jaswal	OK
35	PB168901BS	PB168901BS	LCS	07/18/25 14:06		Jaswal	OK
36	Q2624-03	451	SAM	07/18/25 14:10		Jaswal	OK
37	Q2624-03DUP	451DUP	DUP	07/18/25 14:14		Jaswal	OK
38	Q2624-03L	451L	SD	07/18/25 14:19		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136546

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	CCV03	CCV03	CCV	07/18/25 14:31		Jaswal	OK
40	CCB03	CCB03	CCB	07/18/25 14:36		Jaswal	OK
41	Q2624-03MS	451MS	MS	07/18/25 14:41		Jaswal	OK
42	Q2624-03MSD	451MSD	MSD	07/18/25 14:45		Jaswal	OK
43	Q2624-03A	451A	PS	07/18/25 14:49		Jaswal	OK
44	Q2623-03	3364	SAM	07/18/25 15:26	K,Zn high	Jaswal	Dilution
45	Q2623-03DL	3364DL	SAM	07/18/25 15:30	5x for K,Zn	Jaswal	Confirms
46	PB168913BL	PB168913BL	MB	07/18/25 15:35		Jaswal	OK
47	PB168913BS	PB168913BS	LCS	07/18/25 15:39		Jaswal	OK
48	Q2527-05	ARS20-0015	SAM	07/18/25 15:51	MS/MSD fail for all elements	Jaswal	OK
49	Q2527-05DUP	ARS20-0015DUP	DUP	07/18/25 15:56	MS/MSD fail for all elements	Jaswal	OK
50	Q2527-05L	ARS20-0015L	SD	07/18/25 16:00	MS/MSD fail for all elements	Jaswal	OK
51	CCV04	CCV04	CCV	07/18/25 16:17		Jaswal	OK
52	CCB04	CCB04	CCB	07/18/25 16:21		Jaswal	OK
53	Q2527-05MS	ARS20-0015MS	MS	07/18/25 16:26	MS/MSD fail for all elements	Jaswal	OK
54	Q2527-05MSD	ARS20-0015MSD	MSD	07/18/25 16:30	MS/MSD fail for all elements	Jaswal	OK
55	Q2527-02MSD	ARS20-0015MSD	MSD	07/18/25 16:34	NOT USE	Jaswal	Not Ok
56	Q2527-05A	ARS20-0015A	PS	07/18/25 16:38	MS/MSD fail for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136546

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

57	Q2564-09	ARS20-0039	SAM	07/18/25 16:43		Jaswal	OK
58	Q2595-04	CHRT-24849	SAM	07/18/25 16:47		Jaswal	OK
59	CCV05	CCV05	CCV	07/18/25 16:51	CCV fail for K	Jaswal	OK
60	CCB05	CCB05	CCB	07/18/25 16:56		Jaswal	OK