

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2630

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1000	1000	100	95 - 105	P	07/18/2025	12:01	LB136546
	Iron	3840	4000	96	95 - 105	P	07/18/2025	12:01	LB136546
	Lead	3960	4000	99	95 - 105	P	07/18/2025	12:01	LB136546

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2630

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.9	20.0	109	80 - 120	P	07/18/2025	12:05	LB136546
	Iron	105	100	105	80 - 120	P	07/18/2025	12:05	LB136546
	Lead	12.0	12.0	100	80 - 120	P	07/18/2025	12:05	LB136546

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2630

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	101	90 - 110	P	07/18/2025	12:36	LB136546
	Iron	4980	5000	100	90 - 110	P	07/18/2025	12:36	LB136546
	Lead	5010	5000	100	90 - 110	P	07/18/2025	12:36	LB136546
CCV02	Copper	1270	1250	102	90 - 110	P	07/18/2025	13:29	LB136546
	Iron	5070	5000	102	90 - 110	P	07/18/2025	13:29	LB136546
	Lead	4980	5000	100	90 - 110	P	07/18/2025	13:29	LB136546
CCV03	Copper	1280	1250	102	90 - 110	P	07/18/2025	14:31	LB136546
	Iron	5020	5000	100	90 - 110	P	07/18/2025	14:31	LB136546
	Lead	4960	5000	99	90 - 110	P	07/18/2025	14:31	LB136546
CCV04	Copper	1280	1250	102	90 - 110	P	07/18/2025	16:17	LB136546
	Iron	5220	5000	104	90 - 110	P	07/18/2025	16:17	LB136546
	Lead	4930	5000	99	90 - 110	P	07/18/2025	16:17	LB136546
CCV05	Copper	1280	1250	102	90 - 110	P	07/18/2025	16:51	LB136546
	Iron	5240	5000	105	90 - 110	P	07/18/2025	16:51	LB136546
	Lead	4950	5000	99	90 - 110	P	07/18/2025	16:51	LB136546