

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q2630</u>
Contract:	<u>TULL01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	14.3	2.0	715	-18	22	07/18/2025	12:18	LB136546
	Iron	95900	101000	95	85600	116500	07/18/2025	12:18	LB136546
	Lead	-6.59			-12	12	07/18/2025	12:18	LB136546
ICSAB01	Copper	483	511	94	434	588	07/18/2025	12:23	LB136546
	Iron	96100	99300	97	84400	114500	07/18/2025	12:23	LB136546
	Lead	41.2	49.0	84	37	61	07/18/2025	12:23	LB136546
ICSA	Copper	13.0	2.0	650	-18	22	07/18/2025	12:28	LB136546
	Iron	94300	101000	93	85600	116500	07/18/2025	12:28	LB136546
	Lead	-7.38			-12	12	07/18/2025	12:28	LB136546
ICSAB	Copper	526	511	103	434	588	07/18/2025	12:32	LB136546
	Iron	93400	99300	94	84400	114500	07/18/2025	12:32	LB136546
	Lead	36.3	49.0	74	37	61	07/18/2025	12:32	LB136546