

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2630</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tully Environmental, Inc</u>	Datafile :	<u>VN087361.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0718WBSD01	Benzene	20	18.8	ug/L	94	2		65	135	20
	Toluene	20	18.7	ug/L	94	4		70	130	20
	Ethyl Benzene	20	18.2	ug/L	91	7		60	140	20
	m/p-Xylenes	40	37.0	ug/L	93	6		87	111	20
	o-Xylene	20	18.6	ug/L	93	10		87	111	20