

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2630
 Lab File ID: VN087353.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	46358	7.80	246415	9.08	220107	11.85
UPPER LIMIT	92716	8.3	492830	9.583	440214	12.347
LOWER LIMIT	23179	7.3	123208	8.583	110054	11.347
EPA SAMPLE NO.						
001 Willets Pt Blvd (June)	36588	7.80	185825	9.08	168228	11.85
001 Willets Pt Blvd (June)DL	44931	7.80	209359	9.08	189782	11.85
002 35th Ave (JUNE)	40595	7.79	208925	9.09	187807	11.85
002 35th Ave (JUNE)DL	45506	7.80	211698	9.08	193269	11.85
VN0718WBL01	41937	7.79	219664	9.08	193250	11.85
VN0718WBS01	47522	7.80	244435	9.08	221298	11.85
VN0718WBSD01	42433	7.80	214950	9.08	202696	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.