

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2630
 Lab File ID: VN087353.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001 Willets Pt Blvd (June)	0	0.00				
001 Willets Pt Blvd (June)DL	0	0.00				
002 35th Ave (JUNE)	0	0.00				
002 35th Ave (JUNE)DL	0	0.00				
VN0718WBL01	0	0.00				
VN0718WBS01	0	0.00				
VN0718WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.