

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

SDG NO.: Q2630

Lab File ID: VN087352.D

BFB Injection Date: 07/18/2025

Instrument ID: MSVOA_N

BFB Injection Time: 08:42

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.1
75	30.0 - 60.0% of mass 95	53.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	74.1
175	5.0 - 9.0% of mass 174	5.1 (6.9) 1
176	95.0 - 101.0% of mass 174	70.9 (95.7) 1
177	5.0 - 9.0% of mass 176	4.6 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC020	VSTDCCC020	VN087353.D	07/18/2025	09:15
VN0718WBS01	VN0718WBS01	VN087354.D	07/18/2025	09:49
VN0718WBL01	VN0718WBL01	VN087356.D	07/18/2025	11:04
001 Willets Pt Blvd (June)	Q2630-01	VN087357.D	07/18/2025	11:51
002 35th Ave (JUNE)	Q2630-02	VN087358.D	07/18/2025	12:12
VN0718WBSD01	VN0718WBSD01	VN087361.D	07/18/2025	13:16
001 Willets Pt Blvd (June)DL	Q2630-01DL	VN087362.D	07/18/2025	13:37
002 35th Ave (JUNE)DL	Q2630-02DL	VN087363.D	07/18/2025	13:59