

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN062525

Review By	John Carlone	Review On	6/26/2025 8:30:28 AM
Supervise By	Maresh Dadoda	Supervise On	6/26/2025 2:41:09 PM
SubDirectory	VN062525	HP Acquire Method	HP Processing Method 624N062525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134497		
Initial Calibration Stds	VP134518,VP134519,VP134520,VP134521,VP134522		
CCC	VP134498,VP134524,VP134525		
Internal Standard/PEM			
ICV/I.BLK	VP134523		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087161.D	25 Jun 2025 08:25	JC\MD	Ok
2	VSTDIC005	VN087162.D	25 Jun 2025 08:59	JC\MD	Ok,M
3	VSTDICCC020	VN087163.D	25 Jun 2025 09:20	JC\MD	Ok,M
4	VSTDIC050	VN087164.D	25 Jun 2025 09:41	JC\MD	Ok,M
5	VSTDIC100	VN087165.D	25 Jun 2025 10:03	JC\MD	Ok,M
6	VSTDIC150	VN087166.D	25 Jun 2025 10:24	JC\MD	Ok,M
7	IBLK	VN087167.D	25 Jun 2025 10:46	JC\MD	Ok
8	VSTDICV020	VN087168.D	25 Jun 2025 11:08	JC\MD	Ok,M
9	VN0625WBS01	VN087169.D	25 Jun 2025 11:41	JC\MD	Ok,M
10	VN0625WBSD01	VN087170.D	25 Jun 2025 12:15	JC\MD	Ok,M
11	VN0625WBL01	VN087171.D	25 Jun 2025 12:36	JC\MD	Ok
12	Q2349-01	VN087172.D	25 Jun 2025 13:11	JC\MD	Dilution
13	Q2349-04	VN087173.D	25 Jun 2025 13:33	JC\MD	Dilution
14	Q2349-01DL	VN087174.D	25 Jun 2025 13:54	JC\MD	Ok
15	Q2349-04DL	VN087175.D	25 Jun 2025 14:15	JC\MD	Ok
16	IBLK	VN087176.D	25 Jun 2025 14:37	JC\MD	Ok
17	Q2126-07 2.5PPB	VN087177.D	25 Jun 2025 14:58	JC\MD	Not Ok
18	Q2126-08 5.0PPB	VN087178.D	25 Jun 2025 15:20	JC\MD	Not Ok
19	VSTDCCC020	VN087179.D	25 Jun 2025 16:12	JC\MD	Ok,M

M : Manual Integration