

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN062525

Review By	John Carlone	Review On	6/26/2025 8:30:28 AM
Supervise By	Maresh Dadoda	Supervise On	6/26/2025 2:41:09 PM
SubDirectory	VN062525	HP Acquire Method	HP Processing Method 624N062525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134497
Initial Calibration Stds	VP134518,VP134519,VP134520,VP134521,VP134522
CCC	VP134498,VP134524,VP134525
Internal Standard/PEM	
ICV/I.BLK	VP134523
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087161.D	25 Jun 2025 08:25		JC\MD	Ok
2	VSTDIC005	VSTDIC005	VN087162.D	25 Jun 2025 08:59		JC\MD	Ok,M
3	VSTDICCC020	VSTDICCC020	VN087163.D	25 Jun 2025 09:20		JC\MD	Ok,M
4	VSTDIC050	VSTDIC050	VN087164.D	25 Jun 2025 09:41		JC\MD	Ok,M
5	VSTDIC100	VSTDIC100	VN087165.D	25 Jun 2025 10:03		JC\MD	Ok,M
6	VSTDIC150	VSTDIC150	VN087166.D	25 Jun 2025 10:24		JC\MD	Ok,M
7	IBLK	IBLK	VN087167.D	25 Jun 2025 10:46		JC\MD	Ok
8	VSTDICV020	ICVVN062525	VN087168.D	25 Jun 2025 11:08	pH#Lot#V12668	JC\MD	Ok,M
9	VN0625WBS01	VN0625WBS01	VN087169.D	25 Jun 2025 11:41		JC\MD	Ok,M
10	VN0625WBSD01	VN0625WBSD01	VN087170.D	25 Jun 2025 12:15		JC\MD	Ok,M
11	VN0625WBL01	VN0625WBL01	VN087171.D	25 Jun 2025 12:36		JC\MD	Ok
12	Q2349-01	001-WILLETS-PT-BLVD	VN087172.D	25 Jun 2025 13:11	vial A pH<2 need 20X	JC\MD	Dilution
13	Q2349-04	002-35TH-AVE(JUNE)	VN087173.D	25 Jun 2025 13:33	vial A pH<2 need 20X	JC\MD	Dilution
14	Q2349-01DL	001-WILLETS-PT-BLVD	VN087174.D	25 Jun 2025 13:54	vial B pH<2	JC\MD	Ok
15	Q2349-04DL	002-35TH-AVE(JUNE)	VN087175.D	25 Jun 2025 14:15	vial B pH<2	JC\MD	Ok
16	IBLK	IBLK	VN087176.D	25 Jun 2025 14:37		JC\MD	Ok
17	Q2126-07 2.5PPB	LOD-MDL-WATER-01-C	VN087177.D	25 Jun 2025 14:58	U Flag for com.#06,11	JC\MD	Not Ok
18	Q2126-08 5.0PPB	LOQ-WATER-02-QT2-2	VN087178.D	25 Jun 2025 15:20	comp.#11 recovery is 40%	JC\MD	Not Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN062525

Review By	John Carlone	Review On	6/26/2025 8:30:28 AM
Supervise By	Mahesh Dadoda	Supervise On	6/26/2025 2:41:09 PM
SubDirectory	VN062525	HP Acquire Method	HP Processing Method 624N062525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134497
Initial Calibration Stds	VP134518,VP134519,VP134520,VP134521,VP134522
CCC	VP134498,VP134524,VP134525
Internal Standard/PEM	
ICV/I.BLK	VP134523
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC020	VSTDCCC020EC	VN087179.D	25 Jun 2025 16:12		JC\MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration