

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN071825

Review By	John Carlone	Review On	7/21/2025 8:49:59 AM
Supervise By	Maresh Dadoda	Supervise On	7/21/2025 1:55:50 PM
SubDirectory	VN071825	HP Acquire Method	MSVOA_N HP Processing Method 624n062525w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134817
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134818,VP134819 VP134691

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087352.D	18 Jul 2025 08:42		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087353.D	18 Jul 2025 09:15		JC\MD	Ok,M
3	VN0718WBS01	VN0718WBS01	VN087354.D	18 Jul 2025 09:49		JC\MD	Ok,M
4	VN0718WBS02	VN0718WBS02	VN087355.D	18 Jul 2025 10:21	surrogate failed	JC\MD	Not Ok
5	VN0718WBL01	VN0718WBL01	VN087356.D	18 Jul 2025 11:04		JC\MD	Ok
6	Q2630-01	001 Willets Pt Blvd (Jur	VN087357.D	18 Jul 2025 11:51	vial A pH<2 Need 5X	JC\MD	Dilution
7	Q2630-02	002 35th Ave (JUNE)	VN087358.D	18 Jul 2025 12:12	vial A pH<2 Surrogate Fail;Need 5X	JC\MD	Dilution
8	Q2585-01	EFF-WW	VN087359.D	18 Jul 2025 12:34	vial A pH<2 Surrogate fail;E flag of previous sample	JC\MD	ReRun
9	Q2594-01	CC-071325-RW	VN087360.D	18 Jul 2025 12:55	Surrogate Fail;Run @ lower dilution	JC\MD	Not Ok
10	VN0718WBSD01	VN0718WBSD01	VN087361.D	18 Jul 2025 13:16	BSD Failed Low for com.#16,26	JC\MD	Ok,M
11	Q2630-01DL	001 Willets Pt Blvd (Jur	VN087362.D	18 Jul 2025 13:37	vial B pH<2	JC\MD	Ok
12	Q2630-02DL	002 35th Ave (JUNE)DL	VN087363.D	18 Jul 2025 13:59	vial B pH<2	JC\MD	Ok
13	Q2585-01	EFF-WW	VN087364.D	18 Jul 2025 14:20	vial B pH<2	JC\MD	Ok
14	Q2594-01	CC-071325-RW	VN087365.D	18 Jul 2025 14:41	vial A pH<2	JC\MD	Ok
15	VSTDCCC020	VSTDCCC020EC	VN087366.D	18 Jul 2025 15:02		JC\MD	Ok,M

M : Manual Integration