

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072225

Review By	Semsettin Yesilyurt	Review On	7/23/2025 7:55:47 AM
Supervise By	Maresh Dadoda	Supervise On	7/23/2025 8:20:51 AM
SubDirectory	VL072225	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL07225AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2639
Initial Calibration Stds	AP2641,AP2642,AP2643
CCC	AP2641
Internal Standard/PEM	AP2639
ICV/I.BLK	AP2644
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042722.D	22 Jul 2025 07:25		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042723.D	22 Jul 2025 08:55		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042724.D	22 Jul 2025 09:30		SY/MD	Ok,M
4	VSTDICC001	VSTDICC001	VL042725.D	22 Jul 2025 10:04		SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042726.D	22 Jul 2025 10:38		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042727.D	22 Jul 2025 11:12	Not used	SY/MD	Not Ok
7	VSTDICC0.03	VSTDICC0.03	VL042728.D	22 Jul 2025 11:46		SY/MD	Ok,M
8	VSTDICC015	VSTDICC015	VL042729.D	22 Jul 2025 12:21		SY/MD	Ok,M
9	VSTDICC0.1	VSTDICC0.1	VL042730.D	22 Jul 2025 12:56		SY/MD	Ok,M
10	VSTDICV010	ICVVL072225	VL042731.D	22 Jul 2025 15:06	ICV failed for com.#13,79	SY/MD	Ok,M
11	VL0722ABL01	VL0722ABL01	VL042732.D	22 Jul 2025 15:41		SY/MD	Ok
12	VL0722ABS01	VL0722ABS01	VL042733.D	22 Jul 2025 16:32	BS failed for com.#79	SY/MD	Ok,M
13	Q2617-01	SV1	VL042734.D	22 Jul 2025 17:07		SY/MD	Ok,M
14	Q2617-01DL	SV1DL	VL042735.D	22 Jul 2025 17:42	Not Required	SY/MD	Not Ok
15	Q2617-02	SV2	VL042736.D	22 Jul 2025 18:17		SY/MD	Ok,M
16	Q2617-02DL	SV2DL	VL042737.D	22 Jul 2025 18:51	Not Required	SY/MD	Not Ok
17	Q2617-03	SV3	VL042738.D	22 Jul 2025 19:26	Need 40X	SY/MD	Dilution
18	Q2617-03DL	SV3DL	VL042739.D	22 Jul 2025 20:01		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072225

Review By	Semsettin Yesilyurt	Review On	7/23/2025 7:55:47 AM
Supervise By	Maresh Dadoda	Supervise On	7/23/2025 8:20:51 AM
SubDirectory	VL072225	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL07225AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2639		
Initial Calibration Stds	AP2641,AP2642,AP2643		
CCC	AP2641		
Internal Standard/PEM	AP2639		
ICV/I.BLK	AP2644		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2632-01	IA1	VL042740.D	22 Jul 2025 20:36	need 10X(see non conf.)	SY/MD	Dilution
20	Q2632-01DUP	IA1DUP	VL042741.D	22 Jul 2025 21:12		SY/MD	Ok,M
21	Q2632-01DL	IA1DL	VL042742.D	22 Jul 2025 22:39		SY/MD	Ok,M
22	Q2632-02	OA1	VL042743.D	22 Jul 2025 23:15		SY/MD	Ok,M
23	Q2632-02DL	OA1DL	VL042744.D	22 Jul 2025 23:50	Not Required	SY/MD	Not Ok
24	VL0722ABS02	VL0722ABS02	VL042745.D	23 Jul 2025 00:24	Not Required	SY/MD	Not Ok

M : Manual Integration