

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2632</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL042733.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0722ABS01	Vinyl Chloride	10	8.20	ppbv	82			70	130	
	Heptane	10	9.20	ppbv	92			70	130	
	1,1-Dichloroethene	10	9.30	ppbv	93			70	130	
	Cyclohexane	10	9.50	ppbv	95			70	130	
	cis-1,2-Dichloroethene	10	9.40	ppbv	94			70	130	
	1,1,1-Trichloroethane	10	9.50	ppbv	95			70	130	
	2,2,4-Trimethylpentane	10	9.60	ppbv	96			70	130	
	Benzene	10	9.50	ppbv	95			70	130	
	Trichloroethene	10	9.20	ppbv	92			70	130	
	Toluene	10	9.80	ppbv	98			70	130	
	Tetrachloroethene	10	9.20	ppbv	92			70	130	
	Ethyl Benzene	10	9.60	ppbv	96			70	130	
	m/p-Xylene	20	19.5	ppbv	98			70	130	
	o-Xylene	10	9.70	ppbv	97			70	130	
	1,3,5-Trimethylbenzene	10	10.2	ppbv	102			70	130	
	1,2,4-Trimethylbenzene	10	10.1	ppbv	101			70	130	
	Naphthalene	10	13.6	ppbv	136		*	70	130	
	Hexane	10	9.50	ppbv	95			70	130	