

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2637

Client ID : SSTD020377

Date Analyzed: 08/06/2025

Lab File ID: BN037570.D

Time Analyzed: 13:50

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	68228	7.716	279635	10.50	191629	14.35
UPPER LIMIT	136456	8.216	559270	10.999	383258	14.851
LOWER LIMIT	34114	7.216	139818	9.999	95814.5	13.851
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	65725	7.72	254536	10.50	154893	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.