

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2637

Client ID: SSTD020377

Date Analyzed: 08/06/2025

Lab File ID: BN037570.D

Time Analyzed: 13:50

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	386375	17.086	461792	21.316	518211	24.251
UPPER LIMIT	772750	17.586	923584	21.816	1036420	24.751
LOWER LIMIT	193188	16.586	230896	20.816	259106	23.751
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	310870	17.09	319863	21.32	316534	24.25

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.