



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Nobis Group

SDG No.: Q2638

Project: Raymark Superfund Site

RunNo.: LB136552

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.093	0.099	94	90-110	07/21/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	07/21/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	07/21/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	07/21/2025
Sample ID: Cyanide	CCV4	mg/L	0.25	0.25	100	90-110	07/21/2025