



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.24	0.2	119	70 - 130	CV	07/21/2025	11:59	LB136551
CRI01	Aluminum	107	100	107	65 - 135	P	07/21/2025	13:08	LB136562
	Antimony	54.0	50.0	108	65 - 135	P	07/21/2025	13:08	LB136562
	Arsenic	21.3	20.0	106	65 - 135	P	07/21/2025	13:08	LB136562
	Barium	105	100	105	65 - 135	P	07/21/2025	13:08	LB136562
	Beryllium	6.18	6.0	103	65 - 135	P	07/21/2025	13:08	LB136562
	Cadmium	5.89	6.0	98	65 - 135	P	07/21/2025	13:08	LB136562
	Calcium	2140	2000	107	65 - 135	P	07/21/2025	13:08	LB136562
	Chromium	9.92	10.0	99	65 - 135	P	07/21/2025	13:08	LB136562
	Cobalt	30.8	30.0	102	65 - 135	P	07/21/2025	13:08	LB136562
	Copper	21.8	20.0	109	65 - 135	P	07/21/2025	13:08	LB136562
	Iron	122	100	122	65 - 135	P	07/21/2025	13:08	LB136562
	Lead	12.0	12.0	100	65 - 135	P	07/21/2025	13:08	LB136562
	Magnesium	2220	2000	111	65 - 135	P	07/21/2025	13:08	LB136562
	Manganese	22.8	20.0	114	65 - 135	P	07/21/2025	13:08	LB136562
	Nickel	42.2	40.0	106	65 - 135	P	07/21/2025	13:08	LB136562
	Potassium	1710	2000	85	65 - 135	P	07/21/2025	13:08	LB136562
	Selenium	22.6	20.0	113	65 - 135	P	07/21/2025	13:08	LB136562
	Silver	10.1	10.0	100	65 - 135	P	07/21/2025	13:08	LB136562
	Sodium	1570	2000	79	65 - 135	P	07/21/2025	13:08	LB136562
	Thallium	46.8	40.0	117	65 - 135	P	07/21/2025	13:08	LB136562
	Vanadium	39.0	40.0	97	65 - 135	P	07/21/2025	13:08	LB136562
	Zinc	44.0	40.0	110	65 - 135	P	07/21/2025	13:08	LB136562