

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	93	70 - 130	CV	07/22/2025	15:22	LB136570
CRI	Antimony	1.95	2.0	98	70 - 130	P	07/22/2025	17:12	LB136576
	Arsenic	1.06	1.0	106	70 - 130	P	07/22/2025	17:12	LB136576
	Barium	9.32	10.0	93	70 - 130	P	07/22/2025	17:12	LB136576
	Beryllium	1.20	1.0	120	70 - 130	P	07/22/2025	17:12	LB136576
	Cadmium	1.02	1.0	102	70 - 130	P	07/22/2025	17:12	LB136576
	Chromium	2.02	2.0	101	70 - 130	P	07/22/2025	17:12	LB136576
	Copper	1.65	2.0	82	70 - 130	P	07/22/2025	17:12	LB136576
	Lead	0.84	1.0	84	70 - 130	P	07/22/2025	17:12	LB136576
	Nickel	1.09	1.0	109	70 - 130	P	07/22/2025	17:12	LB136576
	Selenium	5.75	5.0	115	70 - 130	P	07/22/2025	17:12	LB136576
	Silver	1.08	1.0	108	70 - 130	P	07/22/2025	17:12	LB136576
	Thallium	1.03	1.0	103	70 - 130	P	07/22/2025	17:12	LB136576
	Vanadium	4.78	5.0	96	70 - 130	P	07/22/2025	17:12	LB136576
	Zinc	5.34	5.0	107	50 - 150	P	07/22/2025	17:12	LB136576
CRI	Antimony	1.77	2.0	88	70 - 130	P	07/24/2025	12:01	LB136601
	Arsenic	0.98	1.0	98	70 - 130	P	07/24/2025	12:01	LB136601
	Barium	8.47	10.0	85	70 - 130	P	07/24/2025	12:01	LB136601
	Beryllium	1.14	1.0	114	70 - 130	P	07/24/2025	12:01	LB136601
	Cadmium	0.89	1.0	89	70 - 130	P	07/24/2025	12:01	LB136601
	Chromium	1.81	2.0	90	70 - 130	P	07/24/2025	12:01	LB136601
	Copper	1.79	2.0	90	70 - 130	P	07/24/2025	12:01	LB136601
	Lead	0.75	1.0	75	70 - 130	P	07/24/2025	12:01	LB136601
	Nickel	0.95	1.0	95	70 - 130	P	07/24/2025	12:01	LB136601
	Selenium	4.70	5.0	94	70 - 130	P	07/24/2025	12:01	LB136601
	Silver	0.96	1.0	96	70 - 130	P	07/24/2025	12:01	LB136601
	Thallium	0.92	1.0	92	70 - 130	P	07/24/2025	12:01	LB136601
	Vanadium	4.38	5.0	88	70 - 130	P	07/24/2025	12:01	LB136601
	Zinc	4.87	5.0	97	50 - 150	P	07/24/2025	12:01	LB136601