



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB36	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:15	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB34	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:20	LB136570
CCB35	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:59	LB136570
CCB36	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	16:30	LB136570
CCB37	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	17:00	LB136570
CCB38	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	17:33	LB136570
CCB39	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	18:20	LB136570
CCB40	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	18:58	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	16:45	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	16:45	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	16:45	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	16:45	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	16:45	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	16:45	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	16:45	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	16:45	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	16:45	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	16:45	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	16:45	LB136576
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	16:45	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	16:45	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	16:45	LB136576

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	17:09	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	17:09	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	17:09	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	17:09	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	17:09	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	17:09	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	17:09	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	17:09	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	17:09	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	17:09	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	17:09	LB136576
	Thallium	0.090	+/-0.5	J	0.50	1.00	P	07/22/2025	17:09	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	17:09	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	17:09	LB136576
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	18:04	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	18:04	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	18:04	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	18:04	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	18:04	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	18:04	LB136576
	Copper	0.71	+/-1	J	1.50	2.00	P	07/22/2025	18:04	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	18:04	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	18:04	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	18:04	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	18:04	LB136576
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	18:04	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	18:04	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	18:04	LB136576
CCB03	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	19:13	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	19:13	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	19:13	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	19:13	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	19:13	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	19:13	LB136576
	Copper	0.44	+/-1	J	1.50	2.00	P	07/22/2025	19:13	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	19:13	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	19:13	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	19:13	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	19:13	LB136576

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	19:13	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	19:13	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	19:13	LB136576
CCB04	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	20:03	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	20:03	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	20:03	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	20:03	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	20:03	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	20:03	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	20:03	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	20:03	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	20:03	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	20:03	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	20:03	LB136576
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	20:03	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	20:03	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	20:03	LB136576
	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	21:09	LB136576
CCB05	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	21:09	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	21:09	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	21:09	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	21:09	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	21:09	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	21:09	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	21:09	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	21:09	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	21:09	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	21:09	LB136576
	Thallium	0.070	+/-0.5	J	0.50	1.00	P	07/22/2025	21:09	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	21:09	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	21:09	LB136576
	Antimony	0.11	+/-1	U	0.25	2.00	P	07/22/2025	22:11	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	22:11	LB136576
CCB06	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	22:11	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	22:11	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	22:11	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	22:11	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	22:11	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	22:11	LB136576

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	22:11	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	22:11	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	22:11	LB136576
	Thallium	0.080	+/-0.5	J	0.50	1.00	P	07/22/2025	22:11	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	22:11	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	22:11	LB136576
CCB07	Antimony	0.11	+/-1	J	0.25	2.00	P	07/22/2025	22:30	LB136576
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/22/2025	22:30	LB136576
	Barium	0.21	+/-5	U	1.25	10.0	P	07/22/2025	22:30	LB136576
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/22/2025	22:30	LB136576
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/22/2025	22:30	LB136576
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/22/2025	22:30	LB136576
	Copper	0.30	+/-1	U	1.50	2.00	P	07/22/2025	22:30	LB136576
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/22/2025	22:30	LB136576
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/22/2025	22:30	LB136576
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/22/2025	22:30	LB136576
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/22/2025	22:30	LB136576
	Thallium	0.10	+/-0.5	J	0.50	1.00	P	07/22/2025	22:30	LB136576
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/22/2025	22:30	LB136576
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/22/2025	22:30	LB136576

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	07/24/2025	11:38	LB136601
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/24/2025	11:38	LB136601
	Barium	0.21	+/-5	U	1.25	10.0	P	07/24/2025	11:38	LB136601
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/24/2025	11:38	LB136601
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/24/2025	11:38	LB136601
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/24/2025	11:38	LB136601
	Copper	0.30	+/-1	U	1.50	2.00	P	07/24/2025	11:38	LB136601
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/24/2025	11:38	LB136601
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/24/2025	11:38	LB136601
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/24/2025	11:38	LB136601
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	11:38	LB136601
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	11:38	LB136601
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/24/2025	11:38	LB136601
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/24/2025	11:38	LB136601

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.11	+/-1	U	0.25	2.00	P	07/24/2025	11:57	LB136601
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/24/2025	11:57	LB136601
	Barium	0.21	+/-5	U	1.25	10.0	P	07/24/2025	11:57	LB136601
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/24/2025	11:57	LB136601
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/24/2025	11:57	LB136601
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/24/2025	11:57	LB136601
	Copper	0.30	+/-1	U	1.50	2.00	P	07/24/2025	11:57	LB136601
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/24/2025	11:57	LB136601
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/24/2025	11:57	LB136601
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/24/2025	11:57	LB136601
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	11:57	LB136601
	Thallium	0.090	+/-0.5	J	0.50	1.00	P	07/24/2025	11:57	LB136601
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/24/2025	11:57	LB136601
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/24/2025	11:57	LB136601
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	07/24/2025	12:45	LB136601
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/24/2025	12:45	LB136601
	Barium	0.21	+/-5	U	1.25	10.0	P	07/24/2025	12:45	LB136601
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/24/2025	12:45	LB136601
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/24/2025	12:45	LB136601
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/24/2025	12:45	LB136601
	Copper	0.30	+/-1	U	1.50	2.00	P	07/24/2025	12:45	LB136601
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/24/2025	12:45	LB136601
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/24/2025	12:45	LB136601
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/24/2025	12:45	LB136601
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	12:45	LB136601
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	12:45	LB136601
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/24/2025	12:45	LB136601
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/24/2025	12:45	LB136601
CCB03	Antimony	0.11	+/-1	U	0.25	2.00	P	07/24/2025	13:26	LB136601
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/24/2025	13:26	LB136601
	Barium	0.21	+/-5	U	1.25	10.0	P	07/24/2025	13:26	LB136601
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/24/2025	13:26	LB136601
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/24/2025	13:26	LB136601
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/24/2025	13:26	LB136601
	Copper	0.30	+/-1	U	1.50	2.00	P	07/24/2025	13:26	LB136601
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/24/2025	13:26	LB136601
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/24/2025	13:26	LB136601
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/24/2025	13:26	LB136601
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	13:26	LB136601

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	13:26	LB136601
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/24/2025	13:26	LB136601
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/24/2025	13:26	LB136601
CCB04	Antimony	0.11	+/-1	J	0.25	2.00	P	07/24/2025	14:09	LB136601
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	07/24/2025	14:09	LB136601
	Barium	0.21	+/-5	U	1.25	10.0	P	07/24/2025	14:09	LB136601
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	07/24/2025	14:09	LB136601
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	07/24/2025	14:09	LB136601
	Chromium	0.21	+/-1	U	0.75	2.00	P	07/24/2025	14:09	LB136601
	Copper	0.30	+/-1	U	1.50	2.00	P	07/24/2025	14:09	LB136601
	Lead	0.21	+/-0.5	U	0.75	1.00	P	07/24/2025	14:09	LB136601
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	07/24/2025	14:09	LB136601
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	07/24/2025	14:09	LB136601
	Silver	0.060	+/-0.5	U	0.50	1.00	P	07/24/2025	14:09	LB136601
	Thallium	0.080	+/-0.5	J	0.50	1.00	P	07/24/2025	14:09	LB136601
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	07/24/2025	14:09	LB136601
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	07/24/2025	14:09	LB136601