

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168937BS							
Antimony	ug/L	500	511		102	85 - 117	P
Arsenic	ug/L	500	496		99	84 - 116	P
Barium	ug/L	2500	2530		101	86 - 114	P
Beryllium	ug/L	500	505		101	83 - 121	P
Cadmium	ug/L	500	495		99	87 - 115	P
Chromium	ug/L	500	516		103	85 - 116	P
Copper	ug/L	5000	5180		104	85 - 118	P
Lead	ug/L	2500	2500		100	88 - 115	P
Nickel	ug/L	500	506		101	85 - 117	P
Selenium	ug/L	500	503		101	80 - 120	P
Silver	ug/L	500	522		104	85 - 116	P
Thallium	ug/L	500	503		101	82 - 116	P
Vanadium	ug/L	500	511		102	86 - 115	P
Zinc	ug/L	5000	5280		106	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2638

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168959BS Mercury	ug/L	4.0	3.82		96	82 - 119	CV