

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2638

Continuing Calib Date: 07/25/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 03:54

Initial Calibration Time(s): 15:02

16:39

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.51	7.51	7.41	7.61	0.00
2,4-DCAA	7.32	7.33	7.23	7.43	0.01
Dalapon	2.69	2.69	2.59	2.79	0.00
DICHLORPROP	8.22	8.22	8.12	8.32	0.00
2,4-D	8.45	8.46	8.36	8.56	0.01
2,4,5-TP(Silvex)	9.34	9.34	9.24	9.44	0.00
2,4,5-T	9.63	9.64	9.54	9.74	0.01
2,4-DB	10.21	10.21	10.11	10.31	0.00
Dinoseb	11.42	11.43	11.33	11.53	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2638

Continuing Calib Date: 07/25/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 03:54

Initial Calibration Time(s): 15:02

16:39

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
Dalapon	2.71	2.70	2.60	2.80	-0.01
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.02	9.02	8.92	9.12	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.01
2,4-DB	10.92	10.92	10.82	11.02	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00