

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2638

Continuing Calib Date: 07/25/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 07:07

Initial Calibration Time(s): 15:02

16:39

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.51       | 7.51      | 7.41                   | 7.61  | 0.00       |
| 2,4-DCAA         | 7.32       | 7.33      | 7.23                   | 7.43  | 0.01       |
| Dalapon          | 2.69       | 2.69      | 2.59                   | 2.79  | 0.00       |
| DICHLORPROP      | 8.22       | 8.22      | 8.12                   | 8.32  | 0.00       |
| 2,4-D            | 8.45       | 8.46      | 8.36                   | 8.56  | 0.01       |
| 2,4,5-TP(Silvex) | 9.34       | 9.34      | 9.24                   | 9.44  | 0.00       |
| 2,4,5-T          | 9.63       | 9.64      | 9.54                   | 9.74  | 0.01       |
| 2,4-DB           | 10.21      | 10.21     | 10.11                  | 10.31 | 0.00       |
| Dinoseb          | 11.42      | 11.43     | 11.33                  | 11.53 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2638

Continuing Calib Date: 07/25/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 07:07

Initial Calibration Time(s): 15:02

16:39

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.97       | 7.97      | 7.87                   | 8.07  | 0.00       |
| 2,4-DCAA         | 7.77       | 7.77      | 7.67                   | 7.87  | 0.00       |
| Dalapon          | 2.71       | 2.70      | 2.60                   | 2.80  | -0.01      |
| DICHLORPROP      | 8.69       | 8.69      | 8.59                   | 8.79  | 0.00       |
| 2,4-D            | 9.02       | 9.02      | 8.92                   | 9.12  | 0.00       |
| 2,4,5-TP(Silvex) | 9.93       | 9.93      | 9.83                   | 10.03 | 0.00       |
| 2,4,5-T          | 10.35      | 10.36     | 10.26                  | 10.46 | 0.01       |
| 2,4-DB           | 10.92      | 10.92     | 10.82                  | 11.02 | 0.00       |
| Dinoseb          | 11.31      | 11.31     | 11.21                  | 11.41 | 0.00       |