

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2638
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070581.D **Time Analyzed:** 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.487	4.383	4.583	461.200	500.000	-7.8
Aroclor-1016-2	4.506	4.402	4.602	460.220	500.000	-8.0
Aroclor-1016-3	4.564	4.460	4.660	467.240	500.000	-6.6
Aroclor-1016-4	4.653	4.550	4.750	466.570	500.000	-6.7
Aroclor-1016-5	4.938	4.834	5.034	460.670	500.000	-7.9
Aroclor-1260-1	6.027	5.923	6.123	452.340	500.000	-9.5
Aroclor-1260-2	6.283	6.180	6.380	444.670	500.000	-11.1
Aroclor-1260-3	6.634	6.531	6.731	445.720	500.000	-10.9
Aroclor-1260-4	6.848	6.745	6.945	451.190	500.000	-9.8
Aroclor-1260-5	7.154	7.050	7.250	443.530	500.000	-11.3
Decachlorobiphenyl	8.506	8.399	8.599	44.370	50.000	-11.3
Tetrachloro-m-xylene	3.407	3.303	3.503	47.770	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2638
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070581.D **Time Analyzed:** 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.781	3.681	3.881	446.600	500.000	-10.7
Aroclor-1016-2	3.797	3.697	3.897	451.050	500.000	-9.8
Aroclor-1016-3	3.957	3.857	4.057	450.610	500.000	-9.9
Aroclor-1016-4	4.006	3.906	4.106	452.780	500.000	-9.4
Aroclor-1016-5	4.201	4.101	4.301	460.630	500.000	-7.9
Aroclor-1260-1	5.194	5.094	5.294	440.340	500.000	-11.9
Aroclor-1260-2	5.384	5.284	5.484	437.780	500.000	-12.4
Aroclor-1260-3	5.525	5.424	5.624	442.160	500.000	-11.6
Aroclor-1260-4	5.985	5.885	6.085	438.090	500.000	-12.4
Aroclor-1260-5	6.229	6.129	6.329	443.540	500.000	-11.3
Decachlorobiphenyl	7.541	7.440	7.640	43.560	50.000	-12.9
Tetrachloro-m-xylene	2.780	2.680	2.880	46.530	50.000	-6.9